

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX072020\
 Data File : VX017439.D
 Acq On : 20 Jul 2020 17:13
 Operator : JC/SP
 Sample : PB130292TB
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB130292TB

Quant Time: Jul 21 02:40:25 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X071320W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jul 14 02:59:51 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.63	168	241661	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.83	114	426970	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	470371	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.06	152	249923	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.03	65	165701	54.63	ug/l	0.00
Spiked Amount 50.000			Recovery	=	109.26%	
35) Dibromofluoromethane	5.46	113	145476	50.58	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.16%	
50) Toluene-d8	8.70	98	541094	52.16	ug/l	0.00
Spiked Amount 50.000			Recovery	=	104.32%	
62) 4-Bromofluorobenzene	11.12	95	233220	56.59	ug/l	0.00
Spiked Amount 50.000			Recovery	=	113.18%	
Target Compounds						
16) Acetone	2.42	43	33014	30.510	ug/l	100
18) Methyl Acetate	2.75	43	3119	1.167	ug/l #	75
20) Methylene Chloride	2.84	84	34231	12.966	ug/l	97
43) Isopropyl Acetate	6.42	43	84450	13.012	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX072020\
Data File : VX017439.D
Acq On : 20 Jul 2020 17:13
Operator : JC/SP
Sample : PB130292TB
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB130292TB

Quant Time: Jul 21 02:40:25 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X071320W.M
Quant Title : SW846 8260
QLast Update : Tue Jul 14 02:59:51 2020
Response via : Initial Calibration

