

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX072020\
 Data File : VX017443.D
 Acq On : 20 Jul 2020 18:44
 Operator : JC/SP
 Sample : L3177-05
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 AU-05-071720

Quant Time: Jul 21 02:51:53 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X071320W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jul 14 02:59:51 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.64	168	250745	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.83	114	435455	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	466211	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.06	152	252220	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.03	65	170520	54.18	ug/l	0.00
Spiked Amount 50.000			Recovery	=	108.36%	
35) Dibromofluoromethane	5.46	113	149212	50.87	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.74%	
50) Toluene-d8	8.70	98	543187	51.34	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.68%	
62) 4-Bromofluorobenzene	11.12	95	232262	55.26	ug/l	0.00
Spiked Amount 50.000			Recovery	=	110.52%	
Target Compounds						
16) Acetone	2.42	43	10589	9.431	ug/l	98
20) Methylene Chloride	2.84	84	25575	9.336	ug/l	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX072020\
Data File : VX017443.D
Acq On : 20 Jul 2020 18:44
Operator : JC/SP
Sample : L3177-05
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 22 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
AU-05-071720

Quant Time: Jul 21 02:51:53 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X071320W.M
Quant Title : SW846 8260
QLast Update : Tue Jul 14 02:59:51 2020
Response via : Initial Calibration

