

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX072020\
 Data File : VX017444.D
 Acq On : 20 Jul 2020 19:07
 Operator : JC/SP
 Sample : L3177-04
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 AU-05-071720

Quant Time: Jul 21 02:54:19 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X071320W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jul 14 02:59:51 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.63	168	239502	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.84	114	427623	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	475386	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.06	152	254014	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.03	65	168523	56.06	ug/l	0.00
Spiked Amount 50.000			Recovery	=	112.12%	
35) Dibromofluoromethane	5.46	113	145640	50.56	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.12%	
50) Toluene-d8	8.70	98	549751	52.91	ug/l	0.00
Spiked Amount 50.000			Recovery	=	105.82%	
62) 4-Bromofluorobenzene	11.12	95	234828	56.90	ug/l	0.00
Spiked Amount 50.000			Recovery	=	113.80%	
Target Compounds						
16) Acetone	2.42	43	51848	48.348	ug/l	97
18) Methyl Acetate	2.76	43	3442	1.299	ug/l	95
20) Methylene Chloride	2.84	84	49366	18.867	ug/l	99
43) Isopropyl Acetate	6.42	43	74038	11.391	ug/l	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX072020\
Data File : VX017444.D
Acq On : 20 Jul 2020 19:07
Operator : JC/SP
Sample : L3177-04
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 23 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
AU-05-071720

Quant Time: Jul 21 02:54:19 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X071320W.M
Quant Title : SW846 8260
QLast Update : Tue Jul 14 02:59:51 2020
Response via : Initial Calibration

