

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX072020\
 Data File : VX017445.D
 Acq On : 20 Jul 2020 19:30
 Operator : JC/SP
 Sample : L3182-05
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 NB-08-071720

Quant Time: Jul 21 02:57:32 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X071320W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jul 14 02:59:51 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.63	168	225945	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.83	114	401185	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	438936	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.06	152	223407	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.03	65	156042	55.03	ug/l	0.00
Spiked Amount 50.000			Recovery	=	110.06%	
35) Dibromofluoromethane	5.46	113	137104	50.73	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.46%	
50) Toluene-d8	8.70	98	513701	52.70	ug/l	0.00
Spiked Amount 50.000			Recovery	=	105.40%	
62) 4-Bromofluorobenzene	11.12	95	210779	54.43	ug/l	0.00
Spiked Amount 50.000			Recovery	=	108.86%	
Target Compounds						
16) Acetone	2.42	43	40572	40.103	ug/l	98
18) Methyl Acetate	2.75	43	3460	1.385	ug/l #	77
20) Methylene Chloride	2.84	84	50238	20.353	ug/l	98
43) Isopropyl Acetate	6.42	43	64424	10.565	ug/l	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX072020\
Data File : VX017445.D
Acq On : 20 Jul 2020 19:30
Operator : JC/SP
Sample : L3182-05
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 24 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
NB-08-071720

Quant Time: Jul 21 02:57:32 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X071320W.M
Quant Title : SW846 8260
QLast Update : Tue Jul 14 02:59:51 2020
Response via : Initial Calibration

