



Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX072021\
 Data File : VX023356.D
 Acq On : 20 Jul 2021 16:15
 Operator : JC/MD
 Sample : BFB
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X072021W.M
 Title : SW846 8260
 Last Update : Wed Jul 21 06:29:39 2021



AutoFind: Scans 1639, 1640, 1641; Background Corrected with Scan 1632

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	15.3	4699	PASS
75	95	30	60	51.0	15644	PASS
95	95	100	100	100.0	30645	PASS
96	95	5	9	7.1	2167	PASS
173	174	0.00	2	0.0	0	PASS
174	95	50	100	98.4	30149	PASS
175	174	5	9	7.6	2290	PASS
176	174	95	101	97.0	29248	PASS
177	176	5	9	6.9	2027	PASS