

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_X\DATA\VX072118\
 Data File : VX003558.D
 Acq On : 21 Jul 2018 16:40
 Operator : JC/SP
 Sample : J4014-13DL 10X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 GW03-071218DL

Quant Time: Jul 23 02:31:20 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X071918W.M
 Quant Title : SW846 8260
 QLast Update : Thu Jul 19 17:09:44 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	270334	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	446444	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	381958	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	212296	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	180864	56.52	ug/l	0.00
Spiked Amount 50.000			Recovery	=	113.04%	
35) Dibromofluoromethane	5.51	113	153144	45.04	ug/l	0.00
Spiked Amount 50.000			Recovery	=	90.08%	
50) Toluene-d8	8.72	98	580413	45.41	ug/l	0.00
Spiked Amount 50.000			Recovery	=	90.82%	
62) 4-Bromofluorobenzene	11.14	95	198893	42.64	ug/l	0.00
Spiked Amount 50.000			Recovery	=	85.28%	
Target Compounds						
44) Trichloroethene	7.21	130	147872	34.071	ug/l	Qvalue 98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_X\DATA\VX072118\
Data File : VX003558.D
Acq On : 21 Jul 2018 16:40
Operator : JC/SP
Sample : J4014-13DL 10X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
GW03-071218DL

Quant Time: Jul 23 02:31:20 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X071918W.M
Quant Title : SW846 8260
QLast Update : Thu Jul 19 17:09:44 2018
Response via : Initial Calibration

