

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX072118\
 Data File : VX003583.D
 Acq On : 22 Jul 2018 05:01
 Operator : JC/SP
 Sample : J4050-01
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 43 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 KY030SW001-180717

Quant Time: Jul 23 15:14:26 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X071918W.M
 Quant Title : SW846 8260
 QLast Update : Thu Jul 19 17:09:44 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	265189	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	426246	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	377459	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	210213	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	6.07	65	180441	57.48	ug/l	0.00
Spiked Amount			Recovery	=	114.96%	
35) Dibromofluoromethane	5.50	113	163845	50.47	ug/l	0.00
Spiked Amount			Recovery	=	100.94%	
50) Toluene-d8	8.72	98	566251	46.40	ug/l	0.00
Spiked Amount			Recovery	=	92.80%	
62) 4-Bromofluorobenzene	11.14	95	193403	43.43	ug/l	0.00
Spiked Amount			Recovery	=	86.86%	

Target Compounds	Qvalue
------------------	--------

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX072118\
Data File : VX003583.D
Acq On : 22 Jul 2018 05:01
Operator : JC/SP
Sample : J4050-01
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 43 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
KY030SW001-180717

Quant Time: Jul 23 15:14:26 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X071918W.M
Quant Title : SW846 8260
QLast Update : Thu Jul 19 17:09:44 2018
Response via : Initial Calibration

