

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX072118\
 Data File : VX003586.D
 Acq On : 22 Jul 2018 06:21
 Operator : JC/SP
 Sample : J4050-04
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 46 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 KY0300F001-180717

Quant Time: Jul 23 15:15:57 2018
 Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X071918W.M
 Quant Title : SW846 8260
 QLast Update : Thu Jul 19 17:09:44 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	277007	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.87	114	410086	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	384972	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	216663	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	165410	50.45	ug/l	0.00
Spiked Amount	50.000		Recovery	=	100.90%	
35) Dibromofluoromethane	5.51	113	158609	50.78	ug/l	0.00
Spiked Amount	50.000		Recovery	=	101.56%	
50) Toluene-d8	8.72	98	573550	48.85	ug/l	0.00
Spiked Amount	50.000		Recovery	=	97.70%	
62) 4-Bromofluorobenzene	11.14	95	197430	46.08	ug/l	0.00
Spiked Amount	50.000		Recovery	=	92.16%	


Target Compounds	Qvalue


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX072118\
Data File : VX003586.D
Acq On : 22 Jul 2018 06:21
Operator : JC/SP
Sample : J4050-04
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 46 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
KY0300F001-180717

Quant Time: Jul 23 15:15:57 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X071918W.M
Quant Title : SW846 8260
QLast Update : Thu Jul 19 17:09:44 2018
Response via : Initial Calibration





#1

Pentafluorobenzene

Concen: 50.00 ug/l

RT: 5.67 min Scan# 833

Delta R.T. 0.00 min

Lab File: VX003586.D

Acq: 22 Jul 2018 06:21

Instrument :

MSVOA_X

ClientSampleId :


KY0300F001-180717

Tot Ion:168 Resp: 277007

Ion Ratio Lower Upper




168 100

99 48.8 39.3 58.9



Abundance Ion 167.90 (167.60 to 168.60): V

Ion 98.90 (98.60 to 99.60): VX0



#33

1,2-Dichloroethane-d4

Concen: 50.45 ug/l

RT: 6.07 min Scan# 900

Delta R.T. 0.01 min

Lab File: VX003586.D


Acq: 22 Jul 2018 06:21

Tgt Ion: 65 Resp: 165410

Ion Ratio Lower Upper



65 100


67 56.0 0.0 102.6



Abundance Ion 64.90 (64.60 to 65.60): VX0

Ion 66.90 (66.60 to 67.60): VX0





#34

1,4-Difluorobenzene

Concen: 50.00 ug/l

RT: 6.87 min Scan# 1030

Delta R.T. -0.00 min

Lab File: VX003586.D

Acq: 22 Jul 2018 06:21

Instrument :

MSVOA_X

ClientSampleId :

KY0300F001-180717


Tot Ion:114 Resp: 410086

Ion Ratio Lower Upper

114 100

63 18.1 0.0 41.4




88 15.5 0.0 30.0



Abundance Ion 113.90 (113.60 to 114.60): V

Ion 63.00 (62.70 to 63.70): VX0

Ion 87.90 (87.60 to 88.60): VX0



#35

Dibromofluoromethane

Concen: 50.78 ug/l

RT: 5.51 min Scan# 807

Delta R.T. -0.00 min

Lab File: VX003586.D

Acq: 22 Jul 2018 06:21


Tgt Ion:113 Resp: 158609

Ion Ratio Lower Upper

113 100

111 103.3 81.4 122.0



192 25.3 19.1 28.7




Abundance Ion 112.80 (112.50 to 113.50): V

Ion 110.80 (110.50 to 111.50): V

Ion 191.70 (191.40 to 192.40): V









#50
Toluene-d8
Concen: 48.85 ug/l
RT: 8.72 min Scan# 1334
Delta R.T. 0.00 min
Lab File: VX003586.D
Acq: 22 Jul 2018 06:21




Instrument :
MSVOA_X
ClientSampleId :
KY0300F001-180717


Tot Ion: 98 Resp: 573550
Ion Ratio Lower Upper
98 100
100 64.3 50.9 76.3



#62
4-Bromofluorobenzene
Concen: 46.08 ug/l
RT: 11.14 min Scan# 1731
Delta R.T. 0.00 min
Lab File: VX003586.D
Acq: 22 Jul 2018 06:21

Tgt Ion: 95 Resp: 197430
Ion Ratio Lower Upper
95 100
174 99.5 0.0 187.4
176 96.0 0.0 181.0





#63

Chlorobenzene-d5

Concen: 50.00 ug/l

RT: 10.12 min Scan# 1563

Delta R.T. 0.00 min

Lab File: VX003586.D

Acq: 22 Jul 2018 06:21

Instrument :

MSVOA_X

ClientSampleId :

KY0300F001-180717


Tot Ion:117 Resp: 384972

Ion Ratio Lower Upper

117 100

82 51.4 45.0 67.4




119 32.0 25.8 38.6



Abundance Ion 116.90 (116.60 to 117.60): V

Ion 82.00 (81.70 to 82.70): VX0

Ion 118.90 (118.60 to 119.60): V



#72

1,4-Dichlorobenzene-d4

Concen: 50.00 ug/l

RT: 12.08 min Scan# 1885

Delta R.T. 0.00 min

Lab File: VX003586.D

Acq: 22 Jul 2018 06:21


Tgt Ion:152 Resp: 216663

Ion Ratio Lower Upper

152 100

115 53.7 40.1 120.2

150 155.9 0.0 347.2



Abundance Ion 151.90 (151.60 to 152.60): V

Ion 114.90 (114.60 to 115.60): V

Ion 149.90 (149.60 to 150.60): V

