

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX072122\
 Data File : VX030135.D
 Acq On : 20 Jul 2022 16:23
 Operator : JC/MD
 Sample : N3765-01DL 1000X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 SPC-91BN-84DL

Quant Time: Jul 21 03:31:24 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X071222W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jul 13 09:05:23 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.555	168	168018	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	297199	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	253425	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	100616	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	105422	51.549	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery = 103.100%			
35) Dibromofluoromethane	5.397	113	90585	47.711	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery = 95.420%			
50) Toluene-d8	8.652	98	344499	48.245	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery = 96.480%			
62) 4-Bromofluorobenzene	11.085	95	109420	44.157	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery = 88.320%			
Target Compounds						Qvalue
64) Tetrachloroethene	9.274	164	141588	69.599	ug/l	98
90) Hexachloroethane	12.542	117	1158	1.128	ug/l	89

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX072122\
Data File : VX030135.D
Acq On : 20 Jul 2022 16:23
Operator : JC/MD
Sample : N3765-01DL 1000X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
SPC-91BN-84DL

Quant Time: Jul 21 03:31:24 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X071222W.M
Quant Title : SW846 8260
QLast Update : Wed Jul 13 09:05:23 2022
Response via : Initial Calibration

