

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX072122\
 Data File : VX030136.D
 Acq On : 20 Jul 2022 16:46
 Operator : JC/MD
 Sample : N3765-02DL 500X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 SPC-01F-84DL

Quant Time: Jul 21 03:31:45 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X071222W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jul 13 09:05:23 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	166794	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	296312	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	253357	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	97417	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	105228	51.832	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	103.660%
35) Dibromofluoromethane	5.391	113	91417	48.293	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	96.580%
50) Toluene-d8	8.653	98	346990	48.739	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	97.480%
62) 4-Bromofluorobenzene	11.085	95	106692	43.185	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	86.360%
Target Compounds						
64) Tetrachloroethene	9.281	164	86299	42.433	ug/l	Qvalue 97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX072122\
Data File : VX030136.D
Acq On : 20 Jul 2022 16:46
Operator : JC/MD
Sample : N3765-02DL 500X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
SPC-01F-84DL

Quant Time: Jul 21 03:31:45 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X071222W.M
Quant Title : SW846 8260
QLast Update : Wed Jul 13 09:05:23 2022
Response via : Initial Calibration

