

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX072122\
 Data File : VX030137.D
 Acq On : 20 Jul 2022 17:10
 Operator : JC/MD
 Sample : N3765-03 200X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 SPC-01DN-84

Quant Time: Jul 21 03:32:00 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X071222W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jul 13 09:05:23 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	160283	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	290053	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	253034	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	107226	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	104645	53.638	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 107.280%	
35) Dibromofluoromethane	5.391	113	87693	47.326	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 94.660%	
50) Toluene-d8	8.653	98	333405	47.842	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 95.680%	
62) 4-Bromofluorobenzene	11.085	95	112615	46.566	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 93.140%	
Target Compounds						
						Qvalue
44) Trichloroethene	7.135	130	1160	0.527	ug/l	60
64) Tetrachloroethene	9.275	164	254236	125.166	ug/l	98
90) Hexachloroethane	12.536	117	1453	1.328	ug/l	90

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX072122\
Data File : VX030137.D
Acq On : 20 Jul 2022 17:10
Operator : JC/MD
Sample : N3765-03 200X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
SPC-01DN-84

Quant Time: Jul 21 03:32:00 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X071222W.M
Quant Title : SW846 8260
QLast Update : Wed Jul 13 09:05:23 2022
Response via : Initial Calibration

