

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX072122\
 Data File : VX030138.D
 Acq On : 20 Jul 2022 17:34
 Operator : JC/MD
 Sample : N3765-04 200X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 SPC-01A-84

Manual Integrations
 APPROVED

Reviewed By : John Carlone 07/21/2022
 Supervised By : Mahesh Dadoda 07/21/2022

Quant Time: Jul 21 03:32:12 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X071222W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jul 13 09:05:23 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	160839	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	286545	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	240608	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	92457	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	103491	52.863	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery = 105.720%			
35) Dibromofluoromethane	5.391	113	86805	47.420	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery = 94.840%			
50) Toluene-d8	8.653	98	330193	47.961	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery = 95.920%			
62) 4-Bromofluorobenzene	11.085	95	101293	42.397	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery = 84.800%			
Target Compounds						
16) Acetone	2.386	43	1923	2.085	ug/l	88
27) cis-1,2-Dichloroethene	4.507	96	1731m	0.828	ug/l	
64) Tetrachloroethene	9.275	164	22422	11.609	ug/l	95
90) Hexachloroethane	12.542	117	1243	1.318	ug/l	81

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX072122\
Data File : VX030138.D
Acq On : 20 Jul 2022 17:34
Operator : JC/MD
Sample : N3765-04 200X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
SPC-01A-84

Manual Integrations
APPROVED

Reviewed By : John Carlone 07/21/2022
Supervised By : Mahesh Dadoda 07/21/2022

Quant Time: Jul 21 03:32:12 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X071222W.M
Quant Title : SW846 8260
QLast Update : Wed Jul 13 09:05:23 2022
Response via : Initial Calibration

