

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX072122\
 Data File : VX030141.D
 Acq On : 20 Jul 2022 18:44
 Operator : JC/MD
 Sample : N3760-02
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 AUD-1822

Quant Time: Jul 21 03:32:57 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X071222W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jul 13 09:05:23 2022
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	163710	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	295794	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	255002	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	106590	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	106203	53.297	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	106.600%
35) Dibromofluoromethane	5.391	113	90604	47.948	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	95.900%
50) Toluene-d8	8.653	98	346486	48.754	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	97.500%
62) 4-Bromofluorobenzene	11.085	95	110654	44.867	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	89.740%
Target Compounds						Qvalue
16) Acetone	2.386	43	94297	100.428	ug/l	98
17) Carbon Disulfide	2.514	76	2808	0.227	ug/l #	94
25) 2-Butanone	4.562	43	48260	30.173	ug/l	90
44) Trichloroethene	7.141	130	508	0.226	ug/l	85
51) 4-Methyl-2-Pentanone	8.580	43	3593	1.151	ug/l	96
64) Tetrachloroethene	9.275	164	1231	0.601	ug/l #	69
86) p-Isopropyltoluene	12.012	119	1776	0.279	ug/l #	73
95) Naphthalene	13.780	128	3948	0.465	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX072122\
Data File : VX030141.D
Acq On : 20 Jul 2022 18:44
Operator : JC/MD
Sample : N3760-02
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
AUD-1822

Quant Time: Jul 21 03:32:57 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X071222W.M
Quant Title : SW846 8260
QLast Update : Wed Jul 13 09:05:23 2022
Response via : Initial Calibration

