

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX072122\
 Data File : VX030143.D
 Acq On : 20 Jul 2022 19:32
 Operator : JC/MD
 Sample : N3775-01 200X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 M-R-1

Quant Time: Jul 21 03:33:28 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X071222W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jul 13 09:05:23 2022
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	180692	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	327383	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	282058	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	119026	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	115277	52.414	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	104.820%
35) Dibromofluoromethane	5.391	113	97137	46.445	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	92.880%
50) Toluene-d8	8.653	98	376173	47.824	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	95.640%
62) 4-Bromofluorobenzene	11.085	95	123868	45.378	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	90.760%
Target Compounds						
						Qvalue
25) 2-Butanone	4.580	43	3986	2.258	ug/l	93
52) Toluene	8.726	92	26660	4.938	ug/l	97
64) Tetrachloroethene	9.275	164	630	0.278	ug/l #	81
84) 1,2,4-Trimethylbenzene	11.756	105	2027	0.293	ug/l #	57

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX072122\
Data File : VX030143.D
Acq On : 20 Jul 2022 19:32
Operator : JC/MD
Sample : N3775-01 200X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
M-R-1

Quant Time: Jul 21 03:33:28 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X071222W.M
Quant Title : SW846 8260
QLast Update : Wed Jul 13 09:05:23 2022
Response via : Initial Calibration

