

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX072122\
 Data File : VX030144.D
 Acq On : 20 Jul 2022 19:55
 Operator : JC/MD
 Sample : N3775-03 200X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 M-R-2

Quant Time: Jul 21 03:33:39 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X071222W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jul 13 09:05:23 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	181923	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	324247	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	286038	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	122655	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	115245	52.045	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	104.080%
35) Dibromofluoromethane	5.391	113	97703	47.167	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	94.340%
50) Toluene-d8	8.653	98	379004	48.650	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	97.300%
62) 4-Bromofluorobenzene	11.085	95	126208	46.683	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	93.360%
Target Compounds						
64) Tetrachloroethene	9.281	164	515	0.224	ug/l	Qvalue # 73

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX072122\
Data File : VX030144.D
Acq On : 20 Jul 2022 19:55
Operator : JC/MD
Sample : N3775-03 200X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
M-R-2

Quant Time: Jul 21 03:33:39 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X071222W.M
Quant Title : SW846 8260
QLast Update : Wed Jul 13 09:05:23 2022
Response via : Initial Calibration

