

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX072122\
 Data File : VX030147.D
 Acq On : 20 Jul 2022 21:06
 Operator : JC/MD
 Sample : N3767-05 200X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TANK-O

Quant Time: Jul 21 03:34:17 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X071222W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jul 13 09:05:23 2022
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	171097	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	316756	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	273722	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	119555	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	111451	53.516	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	107.040%
35) Dibromofluoromethane	5.391	113	82945	40.990	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	81.980%
50) Toluene-d8	8.653	98	364570	47.904	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	95.800%
62) 4-Bromofluorobenzene	11.085	95	122976	46.563	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	93.120%
Target Compounds						
						Qvalue
16) Acetone	2.386	43	7735	7.882	ug/l	93
44) Trichloroethene	7.129	130	557	0.232	ug/l	97
64) Tetrachloroethene	9.275	164	492	0.224	ug/l #	62
86) p-Isopropyltoluene	12.012	119	2710	0.380	ug/l #	83

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX072122\
Data File : VX030147.D
Acq On : 20 Jul 2022 21:06
Operator : JC/MD
Sample : N3767-05 200X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TANK-O

Quant Time: Jul 21 03:34:17 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X071222W.M
Quant Title : SW846 8260
QLast Update : Wed Jul 13 09:05:23 2022
Response via : Initial Calibration

