

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX072122\
 Data File : VX030150.D
 Acq On : 20 Jul 2022 22:41
 Operator : JC/MD
 Sample : N3781-01 50X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 ORA-0037

Quant Time: Jul 21 03:34:48 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X071222W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jul 13 09:05:23 2022
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	191254	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	341845	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	294721	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	125528	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	120556	51.787	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	103.580%
35) Dibromofluoromethane	5.391	113	103226	47.268	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	94.540%
50) Toluene-d8	8.653	98	393466	47.906	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	95.820%
62) 4-Bromofluorobenzene	11.085	95	133347	46.784	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	93.560%
Target Compounds						
						Qvalue
13) Acrolein	2.252	56	63057	526.971	ug/l	87
64) Tetrachloroethene	9.275	164	604	0.255	ug/l #	70

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX072122\
Data File : VX030150.D
Acq On : 20 Jul 2022 22:41
Operator : JC/MD
Sample : N3781-01 50X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 23 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
ORA-0037

Quant Time: Jul 21 03:34:48 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X071222W.M
Quant Title : SW846 8260
QLast Update : Wed Jul 13 09:05:23 2022
Response via : Initial Calibration

