

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX072122\
 Data File : VX030152.D
 Acq On : 20 Jul 2022 23:28
 Operator : JC/MD
 Sample : N3781-03 50X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 25 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 ORA-0020

Quant Time: Jul 21 03:35:09 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X071222W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jul 13 09:05:23 2022
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	186615	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	335970	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	294938	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	124045	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	121060	53.296	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery	= 106.600%		
35) Dibromofluoromethane	5.391	113	100905	47.013	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery	= 94.020%		
50) Toluene-d8	8.653	98	391488	48.499	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery	= 97.000%		
62) 4-Bromofluorobenzene	11.085	95	130537	46.599	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery	= 93.200%		
Target Compounds						Qvalue
13) Acrolein	2.245	56	788	6.749	ug/l	89
16) Acetone	2.386	43	24733	23.108	ug/l	96
25) 2-Butanone	4.580	43	2242	1.230	ug/l #	78
52) Toluene	8.720	92	11340	2.047	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX072122\
Data File : VX030152.D
Acq On : 20 Jul 2022 23:28
Operator : JC/MD
Sample : N3781-03 50X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 25 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
ORA-0020

Quant Time: Jul 21 03:35:09 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X071222W.M
Quant Title : SW846 8260
QLast Update : Wed Jul 13 09:05:23 2022
Response via : Initial Calibration

