

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX072122\
 Data File : VX030160.D
 Acq On : 21 Jul 2022 14:21
 Operator : JC/MD
 Sample : N3765-04 25X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 SPC-01A-84

Quant Time: Jul 21 15:54:35 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X071222W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jul 13 09:05:23 2022
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	172267	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	313185	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.061	117	272834	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	113553	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	112756	53.775	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 107.560%	
35) Dibromofluoromethane	5.391	113	95339	47.652	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 95.300%	
50) Toluene-d8	8.653	98	364167	48.396	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 96.800%	
62) 4-Bromofluorobenzene	11.085	95	119937	45.930	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 91.860%	
Target Compounds						Qvalue
16) Acetone	2.386	43	1519	1.537	ug/l	92
27) cis-1,2-Dichloroethene	4.495	96	39252	17.521	ug/l	95
30) Chloroform	5.105	83	1573	0.474	ug/l	99
44) Trichloroethene	7.135	130	2932	1.234	ug/l	90
64) Tetrachloroethene	9.281	164	456167	208.282	ug/l	97
90) Hexachloroethane	12.542	117	29057	25.084	ug/l	94
93) 1,2,4-Trichlorobenzene	13.591	180	912	0.400	ug/l	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX072122\
Data File : VX030160.D
Acq On : 21 Jul 2022 14:21
Operator : JC/MD
Sample : N3765-04 25X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 6 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
SPC-01A-84

Quant Time: Jul 21 15:54:35 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X071222W.M
Quant Title : SW846 8260
QLast Update : Wed Jul 13 09:05:23 2022
Response via : Initial Calibration

