

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX072122\
 Data File : VX030161.D
 Acq On : 21 Jul 2022 14:45
 Operator : JC/MD
 Sample : N3765-03DL 1000X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 SPC-01DN-84DL

Quant Time: Jul 21 15:55:09 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X071222W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jul 13 09:05:23 2022
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	167125	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	299798	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	256999	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	105574	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	110826	54.481	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	108.960%
35) Dibromofluoromethane	5.391	113	92110	48.094	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	96.180%
50) Toluene-d8	8.653	98	346222	48.066	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	96.140%
62) 4-Bromofluorobenzene	11.085	95	112506	45.008	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	90.020%
Target Compounds						
						Qvalue
44) Trichloroethene	7.135	130	732	0.322	ug/l	81
64) Tetrachloroethene	9.281	164	47167	22.863	ug/l	94
90) Hexachloroethane	12.536	117	262	0.243	ug/l	89

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX072122\
Data File : VX030161.D
Acq On : 21 Jul 2022 14:45
Operator : JC/MD
Sample : N3765-03DL 1000X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
SPC-01DN-84DL

Quant Time: Jul 21 15:55:09 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X071222W.M
Quant Title : SW846 8260
QLast Update : Wed Jul 13 09:05:23 2022
Response via : Initial Calibration

