

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX072122\
 Data File : VX030163.D
 Acq On : 21 Jul 2022 15:32
 Operator : JC/MD
 Sample : N3765-08DL 200000X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 SPC-04-84DL

Quant Time: Jul 21 16:42:37 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X071222W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jul 13 09:05:23 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	166606	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	301657	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	255277	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	98997	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	109634	54.063	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery = 108.120%			
35) Dibromofluoromethane	5.391	113	88814	46.087	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery = 92.180%			
50) Toluene-d8	8.653	98	348465	48.079	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery = 96.160%			
62) 4-Bromofluorobenzene	11.085	95	108912	43.302	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery = 86.600%			
Target Compounds						
44) Trichloroethene	7.129	130	1181	0.516	ug/l	94
64) Tetrachloroethene	9.281	164	667623	325.797	ug/l	98
90) Hexachloroethane	12.548	117	367	0.363	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX072122\
Data File : VX030163.D
Acq On : 21 Jul 2022 15:32
Operator : JC/MD
Sample : N3765-08DL 200000X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
SPC-04-84DL

Quant Time: Jul 21 16:42:37 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X071222W.M
Quant Title : SW846 8260
QLast Update : Wed Jul 13 09:05:23 2022
Response via : Initial Calibration

