

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX072122\
 Data File : VX030172.D
 Acq On : 21 Jul 2022 19:14
 Operator : JC/MD
 Sample : N3765-09
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 EB-06-07182022

Quant Time: Jul 22 05:00:01 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X071222W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jul 13 09:05:23 2022
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	164087	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	296396	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	260638	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	106980	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	109088	54.619	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	109.240%
35) Dibromofluoromethane	5.391	113	90123	47.596	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	95.200%
50) Toluene-d8	8.653	98	346358	48.637	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	97.280%
62) 4-Bromofluorobenzene	11.085	95	115753	46.839	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	93.680%
Target Compounds						
						Qvalue
16) Acetone	2.392	43	10102	10.734	ug/l	98
25) 2-Butanone	4.574	43	3669	2.289	ug/l #	85
64) Tetrachloroethene	9.269	164	706	0.337	ug/l	85

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX072122\
Data File : VX030172.D
Acq On : 21 Jul 2022 19:14
Operator : JC/MD
Sample : N3765-09
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
EB-06-07182022

Quant Time: Jul 22 05:00:01 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X071222W.M
Quant Title : SW846 8260
QLast Update : Wed Jul 13 09:05:23 2022
Response via : Initial Calibration

