

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX072122\
 Data File : VX030165.D
 Acq On : 21 Jul 2022 16:20
 Operator : JC/MD
 Sample : VX0721WBS03
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :

Quant Time: Jul 21 17:04:52 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X071222W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jul 13 09:05:23 2022
 Response via : Initial Calibration

Manual Integrations APPROVED

Reviewed By :Mahesh Dadoda 07/22/2022
 Supervised By :Semsettin Yesilyurt 07/22/2022

07/22/2022
 Supervised By :Semsettin
 Yesilyurt

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	185249	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	332811	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	299690	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	142770	50.000	ug/l	0.00

System Monitoring Compounds

33) 1,2-Dichloroethane-d4	5.964	65	135438	60.066	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 120.140%	
35) Dibromofluoromethane	5.397	113	109750	51.620	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 103.240%	
50) Toluene-d8	8.653	98	413484	51.710	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 103.420%	
62) 4-Bromofluorobenzene	11.085	95	150966	54.404	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 108.800%	

Target Compounds

						Qvalue
2) Dichlorodifluoromethane	1.166	85	38700	17.756	ug/l	100
3) Chloromethane	1.288	50	38822	20.111	ug/l	96
4) Vinyl Chloride	1.374	62	39005	18.192	ug/l	92
5) Bromomethane	1.599	94	15832	17.382	ug/l	91
6) Chloroethane	1.679	64	22335	18.498	ug/l	96
7) Trichlorofluoromethane	1.880	101	52899	19.104	ug/l	97
8) Diethyl Ether	2.136	74	20165	17.603	ug/l	99
9) 1,1,2-Trichlorotrifluo...	2.325	101	33014	17.464	ug/l	99
10) Methyl Iodide	2.447	142	30162	17.877	ug/l	95
11) Tert butyl alcohol	2.977	59	60231	103.500	ug/l	97
12) 1,1-Dichloroethene	2.319	96	30175	15.988	ug/l	92
13) Acrolein	2.239	56	5685	49.050	ug/l	93
14) Allyl chloride	2.660	41	74766	22.668	ug/l	97
15) Acrylonitrile	3.069	53	154225	116.609	ug/l	99
16) Acetone	2.386	43	128713	121.144	ug/l	97
17) Carbon Disulfide	2.508	76	88651	16.533	ug/l	97
18) Methyl Acetate	2.709	43	76374	23.629	ug/l	98
19) Methyl tert-butyl Ether	3.123	73	131350	22.017	ug/l	95
20) Methylene Chloride	2.788	84	48712	21.605	ug/l	95
21) trans-1,2-Dichloroethene	3.093	96	40909	20.794	ug/l	98
22) Diisopropyl ether	3.770	45	146750	22.730	ug/l	90
23) Vinyl Acetate	3.727	43	645839	118.166	ug/l	99
24) 1,1-Dichloroethane	3.617	63	76880	21.489	ug/l	100
25) 2-Butanone	4.568	43	217882	120.385	ug/l	97
26) 2,2-Dichloropropane	4.483	77	50921	20.262	ug/l	98
27) cis-1,2-Dichloroethene	4.489	96	49713	20.635	ug/l	98
28) Bromochloromethane	4.910	49	35255	23.424	ug/l	95
29) Tetrahydrofuran	5.019	42	144020	117.254	ug/l	97
30) Chloroform	5.099	83	78668	22.065	ug/l	98
31) Cyclohexane	5.477	56	66632	20.385	ug/l	95
32) 1,1,1-Trichloroethane	5.391	97	61721	21.181	ug/l	98
36) 1,1-Dichloropropene	5.696	75	54113	19.704	ug/l	99
37) Ethyl Acetate	4.727	43	77609	21.992	ug/l	98
38) Carbon Tetrachloride	5.684	117	52820	20.625	ug/l	95
39) Methylcyclohexane	7.391	83	65781	18.895	ug/l	99
40) Benzene	6.044	78	176648	20.436	ug/l	97

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX072122\
 Data File : VX030165.D
 Acq On : 21 Jul 2022 16:20
 Operator : JC/MD
 Sample : VX0721WBS03
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :

Quant Time: Jul 21 17:04:52 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X071222W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jul 13 09:05:23 2022
 Response via : Initial Calibration

Manual Integrations APPROVED

Reviewed By :Mahesh Dadoda 07/22/2022
 Supervised By :Semsettin Yesilyurt 07/22/2022

07/22/2022
 Supervised By :Semsettin
 Yesilyurt

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
41) Methacrylonitrile	4.934	41	43492	22.889	ug/l	96
42) 1,2-Dichloroethane	6.098	62	62863	23.466	ug/l	97
43) Isopropyl Acetate	6.355	43	116850	22.020	ug/l	96
44) Trichloroethene	7.129	130	45727	18.104	ug/l	93
45) 1,2-Dichloropropane	7.440	63	48222	21.463	ug/l	97
46) Dibromomethane	7.586	93	31650	20.467	ug/l	97
47) Bromodichloromethane	7.824	83	58737	20.933	ug/l	92
48) Methyl methacrylate	7.702	41	56957	23.263	ug/l	94
49) 1,4-Dioxane	7.671	88	28470	447.554	ug/l	97
51) 4-Methyl-2-Pentanone	8.580	43	414786	118.123	ug/l	97
52) Toluene	8.726	92	109106	19.878	ug/l	100
53) t-1,3-Dichloropropene	8.982	75	58673	20.131	ug/l	98
54) cis-1,3-Dichloropropene	8.372	75	66370	19.877	ug/l	96
55) 1,1,2-Trichloroethane	9.159	97	48431	21.396	ug/l	96
56) Ethyl methacrylate	9.122	69	73326	21.546	ug/l	95
57) 1,3-Dichloropropane	9.311	76	76625	21.310	ug/l	100
58) 2-Chloroethyl Vinyl ether	8.244	63	194874	105.455	ug/l	98
59) 2-Hexanone	9.433	43	326213	121.143	ug/l	96
60) Dibromochloromethane	9.525	129	42637	19.335	ug/l	99
61) 1,2-Dibromoethane	9.616	107	47455	20.144	ug/l	99
64) Tetrachloroethene	9.275	164	48116	20.001	ug/l	98
65) Chlorobenzene	10.086	112	114710	19.222	ug/l	98
66) 1,1,1,2-Tetrachloroethane	10.165	131	40450	19.882	ug/l	99
67) Ethyl Benzene	10.195	91	207393	20.505	ug/l	96
68) m/p-Xylenes	10.305	106	157086	39.598	ug/l	96
69) o-Xylene	10.646	106	77376	19.480	ug/l	94
70) Styrene	10.659	104	132828	20.148	ug/l	99
71) Bromoform	10.805	173	29215	17.874	ug/l #	99
73) Isopropylbenzene	10.963	105	200087	20.127	ug/l	100
74) N-amyl acetate	10.848	43	95255	21.816	ug/l	99
75) 1,1,2,2-Tetrachloroethane	11.213	83	73902	20.605	ug/l	100
76) 1,2,3-Trichloropropane	11.244	75	69157m	21.068	ug/l	
77) Bromobenzene	11.201	156	47887	19.253	ug/l	96
78) n-propylbenzene	11.305	91	237804	20.164	ug/l	100
79) 2-Chlorotoluene	11.366	91	144332	20.432	ug/l	97
80) 1,3,5-Trimethylbenzene	11.457	105	170656	20.373	ug/l	98
81) trans-1,4-Dichloro-2-b...	11.018	75	21276	18.705	ug/l	90
82) 4-Chlorotoluene	11.457	91	161331	20.384	ug/l	100
83) tert-Butylbenzene	11.719	119	158364	19.714	ug/l	95
84) 1,2,4-Trimethylbenzene	11.756	105	170259	20.549	ug/l	98
85) sec-Butylbenzene	11.896	105	216353	20.373	ug/l	99
86) p-Isopropyltoluene	12.012	119	171798	20.165	ug/l	98
87) 1,3-Dichlorobenzene	11.969	146	89186	19.407	ug/l	99
88) 1,4-Dichlorobenzene	12.043	146	88819	18.833	ug/l	99
89) n-Butylbenzene	12.335	91	154418	20.229	ug/l	100
90) Hexachloroethane	12.542	117	27366	18.790	ug/l	97
91) 1,2-Dichlorobenzene	12.335	146	90668	19.419	ug/l	99
92) 1,2-Dibromo-3-Chloropr...	12.945	75	16825	20.545	ug/l	89
93) 1,2,4-Trichlorobenzene	13.591	180	51520	17.965	ug/l	98
94) Hexachlorobutadiene	13.725	225	21131	18.456	ug/l	98
95) Naphthalene	13.780	128	211824	18.609	ug/l	99
96) 1,2,3-Trichlorobenzene	13.963	180	53528	18.344	ug/l	99

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX072122\
Data File : VX030165.D
Acq On : 21 Jul 2022 16:20
Operator : JC/MD
Sample : VX0721WBS03
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 11 Sample Multiplier: 1

Quant Time: Jul 21 17:04:52 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X071222W.M
Quant Title : SW846 8260
QLast Update : Wed Jul 13 09:05:23 2022
Response via : Initial Calibration

Instrument :
MSVOA_X
ClientSampleId :

Manual Integrations
APPROVED

Reviewed By :Mahesh Dadoda 07/22/2022
Supervised By :Semsettin Yesilyurt 07/22/2022

07/22/2022
Supervised By :Semsettin
Yesilyurt

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
----------	------	------	----------	------	-------	----------

07/22/2022

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX072122\
 Data File : VX030165.D
 Acq On : 21 Jul 2022 16:20
 Operator : JC/MD
 Sample : VX0721WBS03
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 11 Sample Multiplier: 1


Instrument :
 MSVOA_X
 ClientSampleId :

Quant Time: Jul 21 17:04:52 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X071222W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jul 13 09:05:23 2022
 Response via : Initial Calibration

Manual Integrations APPROVED

Reviewed By :Mahesh Dadoda 07/22/2022
 Supervised By :Semsettin Yesilyurt 07/22/2022


07/22/2022
 Supervised By :Semsettin
 Yesilyurt





#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 5.556 min Scan# 71
Delta R.T. -0.000 min
Lab File: VX030165.D
Acq: 21 Jul 2022 16:20


Instrument :
MSVOA_X
ClientSampleId :



Tgt Ion:168 Resp: 185249
Ion Ratio Lower Upper
168 100
99 54.7 44.0 66.0



Manual Integrations
APPROVED

Reviewed By :Mahesh Dadoda 07/22/2022
Supervised By :Semsettin Yesilyurt 07/22/2022






07/22/2022
Supervised By :Semsettin
Yesilyurt


07/22/2022



#2
Dichlorodifluoromethane
Concen: 17.756 ug/l
RT: 1.166 min Scan# 13
Delta R.T. 0.000 min
Lab File: VX030165.D
Acq: 21 Jul 2022 16:20


Tgt Ion: 85 Resp: 38700
Ion Ratio Lower Upper
85 100
87 33.1 16.5 49.5





#3
Chloromethane
Concen: 20.111 ug/l
RT: 1.288 min Scan# 31
Delta R.T. 0.000 min
Lab File: VX030165.D
Acq: 21 Jul 2022 16:20



Instrument :
MSVOA_X
ClientSampleId :



Tgt Ion: 50 Resp: 3882
Ion Ratio Lower Upper
50 100
52 30.6 26.3 39.5


Manual Integrations
APPROVED

Reviewed By :Mahesh Dadoda 07/22/2022
Supervised By :Semsettin Yesilyurt 07/22/2022






07/22/2022
Supervised By :Semsettin
Yesilyurt


07/22/2022



#4
Vinyl Chloride
Concen: 18.192 ug/l
RT: 1.374 min Scan# 47
Delta R.T. -0.000 min
Lab File: VX030165.D
Acq: 21 Jul 2022 16:20


Tgt Ion: 62 Resp: 39005
Ion Ratio Lower Upper
62 100
64 36.0 25.2 37.8





#5
Bromomethane
Concen: 17.382 ug/l
RT: 1.599 min Scan# 84
Delta R.T. -0.012 min
Lab File: VX030165.D
Acq: 21 Jul 2022 16:20



Instrument :
MSVOA_X
ClientSampleId :



Tgt Ion: 94 Resp: 1583
Ion Ratio Lower Upper
94 100
96 104.6 76.9 115.3


Manual Integrations
APPROVED

Reviewed By :Mahesh Dadoda 07/22/2022
Supervised By :Semsettin Yesilyurt 07/22/2022




07/22/2022
Supervised By :Semsettin
Yesilyurt



07/22/2022



#6
Chloroethane
Concen: 18.498 ug/l
RT: 1.679 min Scan# 97
Delta R.T. -0.006 min
Lab File: VX030165.D
Acq: 21 Jul 2022 16:20




Tgt Ion: 64 Resp: 22335
Ion Ratio Lower Upper
64 100
66 36.3 27.2 40.8





#7
Trichlorofluoromethane
Concen: 19.104 ug/l
RT: 1.880 min Scan# 11
Delta R.T. -0.006 min
Lab File: VX030165.D
Acq: 21 Jul 2022 16:20



Instrument :
MSVOA_X
ClientSampleId :



Tgt Ion:101 Resp: 5289
Ion Ratio Lower Upper
101 100
103 64.6 49.7 74.5


Manual Integrations
APPROVED

Reviewed By :Mahesh Dadoda 07/22/2022
Supervised By :Semsettin Yesilyurt 07/22/2022




07/22/2022
Supervised By :Semsettin
Yesilyurt



07/22/2022



#8
Diethyl Ether
Concen: 17.603 ug/l
RT: 2.136 min Scan# 172
Delta R.T. -0.000 min
Lab File: VX030165.D
Acq: 21 Jul 2022 16:20




Tgt Ion: 74 Resp: 20165
Ion Ratio Lower Upper
74 100
45 102.1 51.5 154.7





#9
1,1,2-Trichlorotrifluoroethane
Concen: 17.464 ug/l
RT: 2.325 min Scan# 204
Delta R.T. -0.006 min
Lab File: VX030165.D
Acq: 21 Jul 2022 16:20

Instrument :
MSVOA_X
ClientSampleId :






Tgt Ion:101 Resp: 33014
Ion Ratio Lower Upper
101 100
85 44.0 34.0 51.0
151 73.7 59.4 89.0

Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 07/22/2022
Supervised By :Semsettin Yesilyurt 07/22/2022

07/22/2022
Supervised By :Semsettin
Yesilyurt



07/22/2022




#10
Methyl Iodide
Concen: 17.877 ug/l
RT: 2.447 min Scan# 223
Delta R.T. -0.006 min
Lab File: VX030165.D
Acq: 21 Jul 2022 16:20




Tgt Ion:142 Resp: 30162
Ion Ratio Lower Upper
142 100
127 41.5 30.2 45.2
141 14.8 11.3 16.9





#11
Tert butyl alcohol
Concen: 103.500 ug/l
RT: 2.977 min Scan# 313
Delta R.T. -0.018 min
Lab File: VX030165.D
Acq: 21 Jul 2022 16:20

Instrument :
MSVOA_X
ClientSampleId :






Tgt Ion: 59 Resp: 6023
Ion Ratio Lower Upper
59 100
57 10.5 7.4 11.2

Manual Integrations
APPROVED

Reviewed By :Mahesh Dadoda 07/22/2022
Supervised By :Semsettin Yesilyurt 07/22/2022




07/22/2022
Supervised By :Semsettin
Yesilyurt


07/22/2022



#12
1,1-Dichloroethene
Concen: 15.988 ug/l
RT: 2.319 min Scan# 202
Delta R.T. -0.000 min
Lab File: VX030165.D
Acq: 21 Jul 2022 16:20


Tgt Ion: 96 Resp: 30175
Ion Ratio Lower Upper
96 100
61 158.1 117.4 176.0
98 61.5 52.2 78.4





#13
Acrolein
Concen: 49.050 ug/l
RT: 2.239 min Scan# 11
Delta R.T. 0.000 min
Lab File: VX030165.D
Acq: 21 Jul 2022 16:20

Instrument :
MSVOA_X
ClientSampleId :






Tgt Ion: 56 Resp: 568
Ion Ratio Lower Upper
56 100
55 76.1 56.6 84.8

Manual Integrations
APPROVED

Reviewed By :Mahesh Dadoda 07/22/2022
Supervised By :Semsettin Yesilyurt 07/22/2022




07/22/2022
Supervised By :Semsettin
Yesilyurt

07/22/2022



#14
Allyl chloride
Concen: 22.668 ug/l
RT: 2.660 min Scan# 258
Delta R.T. -0.006 min
Lab File: VX030165.D
Acq: 21 Jul 2022 16:20


Tgt Ion: 41 Resp: 74766
Ion Ratio Lower Upper
41 100
39 62.2 48.9 73.3
76 29.9 26.9 40.3





#15
Acrylonitrile
Concen: 116.609 ug/l
RT: 3.069 min Scan# 31
Delta R.T. 0.001 min
Lab File: VX030165.D
Acq: 21 Jul 2022 16:20

Instrument :
MSVOA_X
ClientSampleId :






Tgt Ion: 53 Resp: 154225
Ion Ratio Lower Upper
53 100
52 82.4 65.4 98.0
51 35.6 27.7 41.5

Manual Integrations
APPROVED

Reviewed By :Mahesh Dadoda 07/22/2022
Supervised By :Semsettin Yesilyurt 07/22/2022




07/22/2022
Supervised By :Semsettin
Yesilyurt


07/22/2022



#16
Acetone
Concen: 121.144 ug/l
RT: 2.386 min Scan# 213
Delta R.T. -0.006 min
Lab File: VX030165.D
Acq: 21 Jul 2022 16:20

Tgt Ion: 43 Resp: 128713
Ion Ratio Lower Upper
43 100
58 33.1 27.7 41.5





#17

Carbon Disulfide

Concen: 16.533 ug/l

RT: 2.508 min Scan# 21

Delta R.T. -0.006 min


Lab File: VX030165.D

Acq: 21 Jul 2022 16:20

Instrument :

MSVOA_X

ClientSampleId :



Tgt Ion: 76 Resp: 8865

Ion Ratio Lower Upper

76 100

78 8.6 7.8 11.6

Manual Integrations

APPROVED



Reviewed By :Mahesh Dadoda 07/22/2022

Supervised By :Semsettin Yesilyurt 07/22/2022

07/22/2022

Supervised By :Semsettin
Yesilyurt

07/22/2022



#18

Methyl Acetate

Concen: 23.629 ug/l

RT: 2.709 min Scan# 266

Delta R.T. -0.000 min

Lab File: VX030165.D




Acq: 21 Jul 2022 16:20


Tgt Ion: 43 Resp: 76374

Ion Ratio Lower Upper

43 100


74 23.3 19.4 29.2





#19
Methyl tert-butyl Ether
Concen: 22.017 ug/l
RT: 3.123 min Scan# 31
Delta R.T. 0.006 min
Lab File: VX030165.D
Acq: 21 Jul 2022 16:20

Instrument :
MSVOA_X
ClientSampleId :






Tgt Ion: 73 Resp: 131350
Ion Ratio Lower Upper
73 100
57 21.4 19.0 28.6

Manual Integrations
APPROVED

Reviewed By :Mahesh Dadoda 07/22/2022
Supervised By :Semsettin Yesilyurt 07/22/2022




07/22/2022
Supervised By :Semsettin
Yesilyurt


07/22/2022



#20
Methylene Chloride
Concen: 21.605 ug/l
RT: 2.788 min Scan# 279
Delta R.T. 0.000 min
Lab File: VX030165.D
Acq: 21 Jul 2022 16:20


Tgt Ion: 84 Resp: 48712
Ion Ratio Lower Upper
84 100
49 124.1 93.8 140.6
51 36.7 29.4 44.0
86 60.7 51.0 76.4





#21
trans-1,2-Dichloroethene
Concen: 20.794 ug/l
RT: 3.093 min Scan# 31
Delta R.T. -0.000 min
Lab File: VX030165.D
Acq: 21 Jul 2022 16:20

Instrument :
MSVOA_X
ClientSampleId :






Tgt Ion: 96 Resp: 40909
Ion Ratio Lower Upper
96 100
61 129.3 105.6 158.4
98 63.0 50.1 75.1

Manual Integrations
APPROVED

Reviewed By :Mahesh Dadoda 07/22/2022
Supervised By :Semsettin Yesilyurt 07/22/2022




07/22/2022
Supervised By :Semsettin
Yesilyurt

07/22/2022



#22
Diisopropyl ether
Concen: 22.730 ug/l
RT: 3.770 min Scan# 440
Delta R.T. 0.007 min
Lab File: VX030165.D
Acq: 21 Jul 2022 16:20


Tgt Ion: 45 Resp: 146750
Ion Ratio Lower Upper
45 100
43 69.7 65.4 98.0
87 26.8 22.3 33.5
59 11.4 9.0 13.4





#23
 Vinyl Acetate
 Concen: 118.166 ug/l
 RT: 3.727 min Scan# 41
 Delta R.T. -0.000 min
 Lab File: VX030165.D
 Acq: 21 Jul 2022 16:20

Instrument :
 MSVOA_X
 ClientSampleId :






Tgt Ion: 43 Resp: 645839
 Ion Ratio Lower Upper
 43 100
 86 10.3 8.5 12.7

Manual Integrations
 APPROVED

Reviewed By :Mahesh Dadoda 07/22/2022
 Supervised By :Semsettin Yesilyurt 07/22/2022




07/22/2022
 Supervised By :Semsettin
 Yesilyurt


07/22/2022



#24
 1,1-Dichloroethane
 Concen: 21.489 ug/l
 RT: 3.617 min Scan# 415
 Delta R.T. 0.006 min
 Lab File: VX030165.D
 Acq: 21 Jul 2022 16:20


Tgt Ion: 63 Resp: 76880
 Ion Ratio Lower Upper
 63 100
 98 7.1 3.5 10.5
 100 4.6 2.3 6.9





#25
2-Butanone
Concen: 120.385 ug/l
RT: 4.568 min Scan# 51
Delta R.T. 0.000 min
Lab File: VX030165.D
Acq: 21 Jul 2022 16:20



Instrument :
MSVOA_X
ClientSampleId :



Tgt Ion: 43 Resp: 21788
Ion Ratio Lower Upper
43 100
72 25.4 21.7 32.5


Manual Integrations
APPROVED

Reviewed By :Mahesh Dadoda 07/22/2022
Supervised By :Semsettin Yesilyurt 07/22/2022






07/22/2022
Supervised By :Semsettin
Yesilyurt


07/22/2022



#26
2,2-Dichloropropane
Concen: 20.262 ug/l
RT: 4.483 min Scan# 557
Delta R.T. -0.000 min
Lab File: VX030165.D
Acq: 21 Jul 2022 16:20


Tgt Ion: 77 Resp: 50921
Ion Ratio Lower Upper
77 100
97 24.9 13.0 39.0





#27
 cis-1,2-Dichloroethene
 Concen: 20.635 ug/l
 RT: 4.489 min Scan# 51
 Delta R.T. -0.006 min
 Lab File: VX030165.D
 Acq: 21 Jul 2022 16:20

Instrument :
 MSVOA_X
 ClientSampleId :






Tgt Ion: 96 Resp: 4971
 Ion Ratio Lower Upper
 96 100
 61 137.5 0.0 270.0
 98 62.4 0.0 127.8

Manual Integrations
 APPROVED

Reviewed By :Mahesh Dadoda 07/22/2022
 Supervised By :Semsettin Yesilyurt 07/22/2022




07/22/2022
 Supervised By :Semsettin
 Yesilyurt


07/22/2022



#28
 Bromochloromethane
 Concen: 23.424 ug/l
 RT: 4.910 min Scan# 627
 Delta R.T. 0.007 min
 Lab File: VX030165.D
 Acq: 21 Jul 2022 16:20

Tgt Ion: 49 Resp: 35255
 Ion Ratio Lower Upper
 49 100
 129 1.7 0.0 4.4
 130 74.2 62.8 94.2





#29

Tetrahydrofuran

Concen: 117.254 ug/l

RT: 5.019 min Scan# 64

Delta R.T. 0.006 min


Lab File: VX030165.D

Acq: 21 Jul 2022 16:20

Instrument :

MSVOA_X

ClientSampleId :



Tgt Ion: 42 Resp: 144020

Ion Ratio Lower Upper

42 100

72 43.8 36.5 54.7

71 40.0 33.9 50.9

Manual Integrations

APPROVED



Reviewed By :Mahesh Dadoda 07/22/2022

Supervised By :Semsettin Yesilyurt 07/22/2022

07/22/2022

Supervised By :Semsettin
Yesilyurt

07/22/2022



#30

Chloroform

Concen: 22.065 ug/l

RT: 5.099 min Scan# 658

Delta R.T. -0.006 min

Lab File: VX030165.D




Acq: 21 Jul 2022 16:20


Tgt Ion: 83 Resp: 78668

Ion Ratio Lower Upper

83 100


85 63.1 51.5 77.3





#31
Cyclohexane
Concen: 20.385 ug/l
RT: 5.477 min Scan# 719
Delta R.T. 0.007 min
Lab File: VX030165.D
Acq: 21 Jul 2022 16:20

Instrument :
MSVOA_X
ClientSampleId :






Tgt Ion: 56 Resp: 66632
Ion Ratio Lower Upper
56 100
69 26.8 25.4 38.0
84 83.1 68.9 103.3

Manual Integrations
APPROVED

Reviewed By :Mahesh Dadoda 07/22/2022
Supervised By :Semsettin Yesilyurt 07/22/2022




07/22/2022
Supervised By :Semsettin
Yesilyurt


07/22/2022



#32
1,1,1-Trichloroethane
Concen: 21.181 ug/l
RT: 5.391 min Scan# 706
Delta R.T. 0.000 min
Lab File: VX030165.D
Acq: 21 Jul 2022 16:20

Tgt Ion: 97 Resp: 61721
Ion Ratio Lower Upper
97 100
99 63.4 51.4 77.0
61 46.9 35.8 53.8





#33

1,2-Dichloroethane-d4

Concen: 60.066 ug/l

RT: 5.964 min Scan# 80

Delta R.T. 0.006 min


Lab File: VX030165.D

Acq: 21 Jul 2022 16:20

Instrument :

MSVOA_X

ClientSampleId :



Tgt Ion: 65 Resp: 135438

Ion Ratio Lower Upper

65 100

67 54.1 0.0 109.6

Manual Integrations




APPROVED

Reviewed By :Mahesh Dadoda 07/22/2022
Supervised By :Semsettin Yesilyurt 07/22/2022

07/22/2022

Supervised By :Semsettin
Yesilyurt

07/22/2022



#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 6.769 min Scan# 932

Delta R.T. 0.006 min

Lab File: VX030165.D

Acq: 21 Jul 2022 16:20




Tgt Ion:114 Resp: 332811

Ion Ratio Lower Upper

114 100

63 20.3 0.0 38.4

88 15.3 0.0 31.8





#35

Dibromofluoromethane

Concen: 51.620 ug/l

RT: 5.397 min Scan# 706

Delta R.T. 0.006 min


Lab File: VX030165.D

Acq: 21 Jul 2022 16:20

Instrument :

MSVOA_X

ClientSampleId :



Tgt Ion:	113	Resp:	109750
Ion	Ratio	Lower	Upper
113	100		
111	103.8	83.2	124.8
192	17.1	14.8	22.2

Manual Integrations




APPROVED

Reviewed By :Mahesh Dadoda 07/22/2022
Supervised By :Semsettin Yesilyurt 07/22/2022

07/22/2022

Supervised By :Semsettin
Yesilyurt

07/22/2022



#36

1,1-Dichloropropene

Concen: 19.704 ug/l




RT: 5.696 min Scan# 756


Delta R.T. 0.000 min

Lab File: VX030165.D

Acq: 21 Jul 2022 16:20

Tgt Ion:	75	Resp:	54113
Ion	Ratio	Lower	Upper
75	100		
110	38.5	19.3	57.9
77	31.5	24.4	36.6





#37

Ethyl Acetate

Concen: 21.992 ug/l

RT: 4.727 min Scan# 596

Delta R.T. 0.006 min


Lab File: VX030165.D

Acq: 21 Jul 2022 16:20

Instrument :

MSVOA_X

ClientSampleId :





Tgt Ion: 43	Resp: 77609
Ion Ratio	Lower Upper
43 100	
61 15.1	11.7 17.5
70 10.4	9.3 13.9

Manual Integrations
APPROVED

Reviewed By :Mahesh Dadoda 07/22/2022
Supervised By :Semsettin Yesilyurt 07/22/2022

07/22/2022
Supervised By :Semsettin
Yesilyurt

07/22/2022



#38

Carbon Tetrachloride

Concen: 20.625 ug/l




RT: 5.684 min Scan# 754


Delta R.T. -0.000 min

Lab File: VX030165.D

Acq: 21 Jul 2022 16:20

Tgt Ion:117	Resp: 52820
Ion Ratio	Lower Upper
117 100	
119 97.9	74.1 111.1
121 33.9	25.8 38.8





#39

Methylcyclohexane

Concen: 18.895 ug/l

RT: 7.391 min Scan# 1034

Delta R.T. 0.006 min

Lab File: VX030165.D

Acq: 21 Jul 2022 16:20

Instrument :

MSVOA_X

ClientSampleId :

Tgt Ion: 83 Resp: 6578

Ion	Ratio	Lower	Upper
83	100		
55	72.0	57.7	86.5
98	46.5	36.5	54.7

Manual Integrations





APPROVED

Reviewed By :Mahesh Dadoda 07/22/2022
Supervised By :Semsettin Yesilyurt 07/22/2022

07/22/2022

Supervised By :Semsettin
Yesilyurt

07/22/2022



#40

Benzene

Concen: 20.436 ug/l

RT: 6.044 min Scan# 813




Delta R.T. -0.000 min


Lab File: VX030165.D

Acq: 21 Jul 2022 16:20

Tgt Ion: 78 Resp: 176648


Ion	Ratio	Lower	Upper
78	100		
77	23.8	18.0	27.0





#41
Methacrylonitrile
Concen: 22.889 ug/l
RT: 4.934 min Scan# 61
Delta R.T. 0.006 min
Lab File: VX030165.D
Acq: 21 Jul 2022 16:20

Instrument :
MSVOA_X
ClientSampleId :






Tgt Ion: 41 Resp: 43492
Ion Ratio Lower Upper
41 100
39 55.8 42.9 64.3
67 70.9 60.1 90.1
52 29.4 22.6 34.0

Manual Integrations
APPROVED

Reviewed By :Mahesh Dadoda 07/22/2022
Supervised By :Semsettin Yesilyurt 07/22/2022




07/22/2022
Supervised By :Semsettin
Yesilyurt


07/22/2022



#42
1,2-Dichloroethane
Concen: 23.466 ug/l
RT: 6.098 min Scan# 822
Delta R.T. 0.006 min
Lab File: VX030165.D
Acq: 21 Jul 2022 16:20

Tgt Ion: 62 Resp: 62863
Ion Ratio Lower Upper
62 100
98 9.0 0.0 20.2





#43

Isopropyl Acetate

Concen: 22.020 ug/l

RT: 6.355 min Scan# 80

Delta R.T. 0.013 min


Lab File: VX030165.D

Acq: 21 Jul 2022 16:20

Instrument :

MSVOA_X

ClientSampleId :



Tgt Ion: 43	Resp: 11685
Ion Ratio	Lower Upper
43 100	
61 18.9	16.5 24.7
87 12.5	11.1 16.7

Manual Integrations



APPROVED

Reviewed By :Mahesh Dadoda 07/22/2022
Supervised By :Semsettin Yesilyurt 07/22/2022

07/22/2022

Supervised By :Semsettin Yesilyurt

07/22/2022



#44

Trichloroethene

Concen: 18.104 ug/l




RT: 7.129 min Scan# 991


Delta R.T. -0.000 min

Lab File: VX030165.D

Acq: 21 Jul 2022 16:20


Tgt Ion:130	Resp:	45727	
Ion	Ratio	Lower	Upper
130	100		
95	104.7	0.0	194.8





#45
1,2-Dichloropropane
Concen: 21.463 ug/l
RT: 7.440 min Scan# 1041
Delta R.T. 0.006 min
Lab File: VX030165.D
Acq: 21 Jul 2022 16:20



Instrument :
MSVOA_X
ClientSampleId :



Tgt Ion: 63 Resp: 4822
Ion Ratio Lower Upper
63 100
65 29.8 25.2 37.8


Manual Integrations
APPROVED

Reviewed By :Mahesh Dadoda 07/22/2022
Supervised By :Semsettin Yesilyurt 07/22/2022






07/22/2022
Supervised By :Semsettin
Yesilyurt


07/22/2022



#46
Dibromomethane
Concen: 20.467 ug/l
RT: 7.586 min Scan# 1066
Delta R.T. 0.000 min
Lab File: VX030165.D
Acq: 21 Jul 2022 16:20


Tgt Ion: 93 Resp: 31650
Ion Ratio Lower Upper
93 100
95 82.7 67.4 101.0
174 91.3 75.6 113.4





#47
Bromodichloromethane
Concen: 20.933 ug/l
RT: 7.824 min Scan# 1105
Delta R.T. -0.000 min
Lab File: VX030165.D
Acq: 21 Jul 2022 16:20

Instrument :
MSVOA_X
ClientSampleId :






Tgt Ion: 83 Resp: 5873
Ion Ratio Lower Upper
83 100
85 60.0 53.5 80.3
127 7.4 6.6 10.0

Manual Integrations
APPROVED

Reviewed By :Mahesh Dadoda 07/22/2022
Supervised By :Semsettin Yesilyurt 07/22/2022




07/22/2022
Supervised By :Semsettin
Yesilyurt

07/22/2022



#48
Methyl methacrylate
Concen: 23.263 ug/l
RT: 7.702 min Scan# 1085
Delta R.T. 0.006 min
Lab File: VX030165.D
Acq: 21 Jul 2022 16:20

Tgt Ion: 41 Resp: 56957
Ion Ratio Lower Upper
41 100
69 80.6 68.8 103.2
39 60.2 44.6 66.8





#49

1,4-Dioxane

Concen: 447.554 ug/l

RT: 7.671 min Scan# 1080

Delta R.T. -0.012 min


Lab File: VX030165.D

Acq: 21 Jul 2022 16:20

Instrument :

MSVOA_X

ClientSampleId :






Tgt Ion:	88	Resp:	28470
Ion Ratio	Lower	Upper	
88	100		
43	33.7	23.8	35.8
58	68.6	53.8	80.6

Manual Integrations
APPROVED

Reviewed By :Mahesh Dadoda 07/22/2022
Supervised By :Semsettin Yesilyurt 07/22/2022

07/22/2022
Supervised By :Semsettin
Yesilyurt

07/22/2022



#50

Toluene-d8

Concen: 51.710 ug/l




RT: 8.653 min Scan# 1241

Delta R.T. -0.000 min

Lab File: VX030165.D

Acq: 21 Jul 2022 16:20


Tgt Ion:	98	Resp:	413484
Ion Ratio	Lower	Upper	
98	100		
100	64.5	52.3	78.5





#51
4-Methyl-2-Pentanone
Concen: 118.123 ug/l
RT: 8.580 min Scan# 1229
Delta R.T. -0.000 min
Lab File: VX030165.D
Acq: 21 Jul 2022 16:20

Instrument :
MSVOA_X
ClientSampleId :






Tgt Ion: 43 Resp: 414786
Ion Ratio Lower Upper
43 100
58 37.3 31.4 47.2

Manual Integrations
APPROVED

Reviewed By :Mahesh Dadoda 07/22/2022
Supervised By :Semsettin Yesilyurt 07/22/2022




07/22/2022
Supervised By :Semsettin
Yesilyurt


07/22/2022



#52
Toluene
Concen: 19.878 ug/l
RT: 8.726 min Scan# 1253
Delta R.T. 0.006 min
Lab File: VX030165.D
Acq: 21 Jul 2022 16:20

Tgt Ion: 92 Resp: 109106
Ion Ratio Lower Upper
92 100
91 173.4 139.2 208.8





#53

t-1,3-Dichloropropene

Concen: 20.131 ug/l

RT: 8.982 min Scan# 1295

Delta R.T. 0.000 min

Lab File: VX030165.D

Acq: 21 Jul 2022 16:20

Instrument :

MSVOA_X

ClientSampleId :

Tgt Ion: 75 Resp: 5867

Ion Ratio Lower Upper

75 100

77 33.8 26.3 39.5

Manual Integrations

APPROVED





Reviewed By :Mahesh Dadoda 07/22/2022

Supervised By :Semsettin Yesilyurt 07/22/2022

07/22/2022

Supervised By :Semsettin
Yesilyurt

07/22/2022



#54

cis-1,3-Dichloropropene

Concen: 19.877 ug/l

RT: 8.372 min Scan# 1195

Delta R.T. 0.000 min

Lab File: VX030165.D

Acq: 21 Jul 2022 16:20




Tgt Ion: 75 Resp: 66370


Ion Ratio Lower Upper

75 100

77 30.6 25.0 37.4


39 50.4 37.5 56.3





#55
1,1,2-Trichloroethane
Concen: 21.396 ug/l
RT: 9.159 min Scan# 1318
Delta R.T. 0.006 min
Lab File: VX030165.D
Acq: 21 Jul 2022 16:20

Instrument :
MSVOA_X
ClientSampleId :






Tgt Ion: 97 Resp: 4843
Ion Ratio Lower Upper
97 100
83 80.4 68.9 103.3
85 51.5 44.0 66.0
99 59.9 48.3 72.5

Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 07/22/2022
Supervised By :Semsettin Yesilyurt 07/22/2022

07/22/2022
Supervised By :Semsettin
Yesilyurt



07/22/2022




#56
Ethyl methacrylate
Concen: 21.546 ug/l
RT: 9.122 min Scan# 1318
Delta R.T. 0.006 min
Lab File: VX030165.D
Acq: 21 Jul 2022 16:20



Tgt Ion: 69 Resp: 73326
Ion Ratio Lower Upper
69 100
41 69.0 52.3 78.5
39 38.7 28.2 42.4





#57

1,3-Dichloropropane

Concen: 21.310 ug/l

RT: 9.311 min Scan# 11

Delta R.T. 0.000 min


Lab File: VX030165.D

Acq: 21 Jul 2022 16:20

Instrument :

MSVOA_X

ClientSampleId :



Tgt Ion: 76 Resp: 7662

Ion Ratio Lower Upper

76 100

78 32.3 25.9 38.9

Manual Integrations



APPROVED

Reviewed By :Mahesh Dadoda 07/22/2022
Supervised By :Semsettin Yesilyurt 07/22/2022

07/22/2022

Supervised By :Semsettin
Yesilyurt

07/22/2022



#58

2-Chloroethyl Vinyl ether

Concen: 105.455 ug/l

RT: 8.244 min Scan# 1174

Delta R.T. 0.000 min

Lab File: VX030165.D




Acq: 21 Jul 2022 16:20


Tgt Ion: 63 Resp: 194874

Ion Ratio Lower Upper

63 100


106 26.4 22.2 33.2





#59
2-Hexanone
Concen: 121.143 ug/l
RT: 9.433 min Scan# 1369
Delta R.T. 0.000 min
Lab File: VX030165.D
Acq: 21 Jul 2022 16:20


Instrument :
MSVOA_X
ClientSampleId :



Tgt Ion: 43 Resp: 32621
Ion Ratio Lower Upper
43 100
58 52.6 27.6 82.8



Manual Integrations
APPROVED

Reviewed By :Mahesh Dadoda 07/22/2022
Supervised By :Semsettin Yesilyurt 07/22/2022






07/22/2022
Supervised By :Semsettin
Yesilyurt

07/22/2022



#60
Dibromochloromethane
Concen: 19.335 ug/l
RT: 9.525 min Scan# 1384
Delta R.T. -0.000 min
Lab File: VX030165.D
Acq: 21 Jul 2022 16:20


Tgt Ion:129 Resp: 42637
Ion Ratio Lower Upper
129 100
127 78.0 38.5 115.5





#61
1,2-Dibromoethane
Concen: 20.144 ug/l
RT: 9.616 min Scan# 11
Delta R.T. 0.006 min
Lab File: VX030165.D
Acq: 21 Jul 2022 16:20

Instrument :
MSVOA_X
ClientSampleId :





Tgt Ion:107 Resp: 4745
Ion Ratio Lower Upper
107 100
109 94.9 76.6 114.8

Manual Integrations
APPROVED

Reviewed By :Mahesh Dadoda 07/22/2022
Supervised By :Semsettin Yesilyurt 07/22/2022




07/22/2022
Supervised By :Semsettin
Yesilyurt

07/22/2022



#62
4-Bromofluorobenzene
Concen: 54.404 ug/l
RT: 11.085 min Scan# 1640
Delta R.T. 0.006 min
Lab File: VX030165.D
Acq: 21 Jul 2022 16:20


Tgt Ion: 95 Resp: 150966
Ion Ratio Lower Upper
95 100
174 71.5 0.0 149.0
176 70.5 0.0 146.0





#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 10.055 min Scan# 1471
Delta R.T. 0.000 min
Lab File: VX030165.D
Acq: 21 Jul 2022 16:20

Instrument :
MSVOA_X
ClientSampleId :






Tgt Ion:117 Resp: 299690
Ion Ratio Lower Upper
117 100
82 59.2 42.7 64.1
119 31.4 25.4 38.2

Manual Integrations
APPROVED

Reviewed By :Mahesh Dadoda 07/22/2022
Supervised By :Semsettin Yesilyurt 07/22/2022




07/22/2022
Supervised By :Semsettin
Yesilyurt


07/22/2022



#64
Tetrachloroethene
Concen: 20.001 ug/l
RT: 9.275 min Scan# 1343
Delta R.T. -0.000 min
Lab File: VX030165.D
Acq: 21 Jul 2022 16:20


Tgt Ion:164 Resp: 48116
Ion Ratio Lower Upper
164 100
166 118.9 98.2 147.4
129 98.6 80.8 121.2
131 93.2 75.5 113.3





#65
Chlorobenzene
Concen: 19.222 ug/l
RT: 10.086 min Scan# 1475
Delta R.T. 0.007 min
Lab File: VX030165.D
Acq: 21 Jul 2022 16:20


Instrument :
MSVOA_X
ClientSampleId :



Tgt Ion:112 Resp: 114710
Ion Ratio Lower Upper
112 100
114 29.6 24.6 36.8



Manual Integrations
APPROVED

Reviewed By :Mahesh Dadoda 07/22/2022
Supervised By :Semsettin Yesilyurt 07/22/2022






07/22/2022
Supervised By :Semsettin
Yesilyurt


07/22/2022



#66
1,1,1,2-Tetrachloroethane
Concen: 19.882 ug/l
RT: 10.165 min Scan# 1489
Delta R.T. -0.000 min
Lab File: VX030165.D
Acq: 21 Jul 2022 16:20


Tgt Ion:131 Resp: 40450
Ion Ratio Lower Upper
131 100
133 94.6 46.9 140.8
119 68.9 34.4 103.1





#67
Ethyl Benzene
Concen: 20.505 ug/l
RT: 10.195 min Scan# 1494
Delta R.T. 0.000 min
Lab File: VX030165.D
Acq: 21 Jul 2022 16:20



Instrument :
MSVOA_X
ClientSampleId :



Tgt Ion: 91 Resp: 20739
Ion Ratio Lower Upper
91 100
106 29.8 25.5 38.3


Manual Integrations
APPROVED

Reviewed By :Mahesh Dadoda 07/22/2022
Supervised By :Semsettin Yesilyurt 07/22/2022






07/22/2022
Supervised By :Semsettin
Yesilyurt

07/22/2022



#68
m/p-Xylenes
Concen: 39.598 ug/l
RT: 10.305 min Scan# 1512
Delta R.T. 0.000 min
Lab File: VX030165.D
Acq: 21 Jul 2022 16:20


Tgt Ion:106 Resp: 157086
Ion Ratio Lower Upper
106 100
91 204.5 158.2 237.4





#69
o-Xylene
Concen: 19.480 ug/l
RT: 10.646 min Scan# 1568
Delta R.T. 0.000 min
Lab File: VX030165.D
Acq: 21 Jul 2022 16:20



Instrument :
MSVOA_X
ClientSampleId :



Tgt Ion:106 Resp: 77370
Ion Ratio Lower Upper
106 100
91 216.7 103.8 311.6


Manual Integrations
APPROVED

Reviewed By :Mahesh Dadoda 07/22/2022
Supervised By :Semsettin Yesilyurt 07/22/2022




07/22/2022
Supervised By :Semsettin
Yesilyurt



07/22/2022




#70
Styrene
Concen: 20.148 ug/l
RT: 10.659 min Scan# 1570
Delta R.T. -0.000 min
Lab File: VX030165.D
Acq: 21 Jul 2022 16:20




Tgt Ion:104 Resp: 132828
Ion Ratio Lower Upper
104 100
78 48.9 37.8 56.8
103 55.2 44.5 66.7





#71
Bromoform
Concen: 17.874 ug/l
RT: 10.805 min Scan# 1594
Delta R.T. -0.000 min
Lab File: VX030165.D
Acq: 21 Jul 2022 16:20

Instrument :
MSVOA_X
ClientSampleId :






Tgt Ion:173 Resp: 29215
Ion Ratio Lower Upper
173 100
175 48.5 23.8 71.2
254 0.0 0.1 0.1

Manual Integrations
APPROVED

Reviewed By :Mahesh Dadoda 07/22/2022
Supervised By :Semsettin Yesilyurt 07/22/2022




07/22/2022
Supervised By :Semsettin
Yesilyurt


07/22/2022



#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 12.024 min Scan# 1794
Delta R.T. 0.000 min
Lab File: VX030165.D
Acq: 21 Jul 2022 16:20


Tgt Ion:152 Resp: 142770
Ion Ratio Lower Upper
152 100
115 70.8 40.5 121.5
150 164.3 0.0 347.0





#73
Isopropylbenzene
Concen: 20.127 ug/l
RT: 10.963 min Scan# 1620
Delta R.T. 0.000 min
Lab File: VX030165.D
Acq: 21 Jul 2022 16:20


Instrument :
MSVOA_X
ClientSampleId :



Tgt Ion: 105 Resp: 20008
Ion Ratio Lower Upper
105 100
120 26.8 13.5 40.4



Manual Integrations
APPROVED

Reviewed By :Mahesh Dadoda 07/22/2022
Supervised By :Semsettin Yesilyurt 07/22/2022






07/22/2022
Supervised By :Semsettin
Yesilyurt


07/22/2022



#74
N-ethyl acetate
Concen: 21.816 ug/l
RT: 10.848 min Scan# 1601
Delta R.T. -0.000 min
Lab File: VX030165.D
Acq: 21 Jul 2022 16:20


Tgt Ion: 43 Resp: 95255
Ion Ratio Lower Upper
43 100
70 40.4 32.3 48.5
55 24.1 18.4 27.6
61 22.9 19.2 28.8





#75
1,1,2,2-Tetrachloroethane
Concen: 20.605 ug/l
RT: 11.213 min Scan# 1661
Delta R.T. 0.000 min
Lab File: VX030165.D
Acq: 21 Jul 2022 16:20

Instrument :
MSVOA_X
ClientSampleId :






Tgt Ion: 83 Resp: 73902
Ion Ratio Lower Upper
83 100
131 9.3 4.5 13.7
85 65.3 32.5 97.5

Manual Integrations
APPROVED

Reviewed By :Mahesh Dadoda 07/22/2022
Supervised By :Semsettin Yesilyurt 07/22/2022




07/22/2022
Supervised By :Semsettin
Yesilyurt


07/22/2022



#76
1,2,3-Trichloropropane
Concen: 21.068 ug/l m
RT: 11.244 min Scan# 1666
Delta R.T. -0.000 min
Lab File: VX030165.D
Acq: 21 Jul 2022 16:20


Tgt Ion: 75 Resp: 69157
Ion Ratio Lower Upper
75 100
77 39.2 19.8 59.3





#77
Bromobenzene
Concen: 19.253 ug/l
RT: 11.201 min Scan# 1659
Delta R.T. 0.000 min
Lab File: VX030165.D
Acq: 21 Jul 2022 16:20

Instrument :
MSVOA_X
ClientSampleId :






Tgt Ion:156 Resp: 4788
Ion Ratio Lower Upper
156 100
77 172.7 82.7 247.9
158 94.0 47.5 142.5

Manual Integrations
APPROVED

Reviewed By :Mahesh Dadoda 07/22/2022
Supervised By :Semsettin Yesilyurt 07/22/2022




07/22/2022
Supervised By :Semsettin
Yesilyurt

07/22/2022



#78
n-propylbenzene
Concen: 20.164 ug/l
RT: 11.305 min Scan# 1676
Delta R.T. -0.000 min
Lab File: VX030165.D
Acq: 21 Jul 2022 16:20


Tgt Ion: 91 Resp: 237804
Ion Ratio Lower Upper
91 100
120 23.0 11.6 34.8





#79
2-Chlorotoluene
Concen: 20.432 ug/l
RT: 11.366 min Scan# 1101
Delta R.T. -0.000 min
Lab File: VX030165.D
Acq: 21 Jul 2022 16:20



Instrument :
MSVOA_X
ClientSampleId :



Tgt Ion: 91 Resp: 144332
Ion Ratio Lower Upper
91 100
126 33.3 17.4 52.3


Manual Integrations
APPROVED

Reviewed By :Mahesh Dadoda 07/22/2022
Supervised By :Semsettin Yesilyurt 07/22/2022




07/22/2022
Supervised By :Semsettin
Yesilyurt



07/22/2022




#80
1,3,5-Trimethylbenzene
Concen: 20.373 ug/l
RT: 11.457 min Scan# 1701
Delta R.T. 0.006 min
Lab File: VX030165.D
Acq: 21 Jul 2022 16:20




Tgt Ion:105 Resp: 170656
Ion Ratio Lower Upper
105 100
120 47.4 24.4 73.2





#81
trans-1,4-Dichloro-2-butene
Concen: 18.705 ug/l
RT: 11.018 min Scan# 1629
Delta R.T. 0.000 min
Lab File: VX030165.D
Acq: 21 Jul 2022 16:20



Instrument :
MSVOA_X
ClientSampleId :



Tgt Ion: 75 Resp: 21270
Ion Ratio Lower Upper
75 100
53 107.0 74.2 111.4
89 44.0 34.8 52.2


Manual Integrations
APPROVED

Reviewed By :Mahesh Dadoda 07/22/2022
Supervised By :Semsettin Yesilyurt 07/22/2022




07/22/2022
Supervised By :Semsettin
Yesilyurt



07/22/2022




#82
4-Chlorotoluene
Concen: 20.384 ug/l
RT: 11.457 min Scan# 1701
Delta R.T. 0.000 min
Lab File: VX030165.D
Acq: 21 Jul 2022 16:20




Tgt Ion: 91 Resp: 161331
Ion Ratio Lower Upper
91 100
126 30.5 15.3 45.9





#83
tert-Butylbenzene
Concen: 19.714 ug/l
RT: 11.719 min Scan# 1744
Delta R.T. 0.000 min
Lab File: VX030165.D
Acq: 21 Jul 2022 16:20

Instrument :
MSVOA_X
ClientSampleId :






Tgt Ion:119 Resp: 158364
Ion Ratio Lower Upper
119 100
91 61.3 28.3 84.9
134 23.3 11.2 33.5

Manual Integrations
APPROVED

Reviewed By :Mahesh Dadoda 07/22/2022
Supervised By :Semsettin Yesilyurt 07/22/2022




07/22/2022
Supervised By :Semsettin
Yesilyurt

07/22/2022



#84
1,2,4-Trimethylbenzene
Concen: 20.549 ug/l
RT: 11.756 min Scan# 1750
Delta R.T. 0.000 min
Lab File: VX030165.D
Acq: 21 Jul 2022 16:20


Tgt Ion:105 Resp: 170259
Ion Ratio Lower Upper
105 100
120 44.4 22.9 68.8





#85
 sec-Butylbenzene
 Concen: 20.373 ug/l
 RT: 11.896 min Scan# 11
 Delta R.T. 0.006 min
 Lab File: VX030165.D
 Acq: 21 Jul 2022 16:20

Instrument :
 MSVOA_X
 ClientSampleId :






Tgt Ion:105 Resp: 21635
 Ion Ratio Lower Upper
 105 100
 134 19.5 10.0 30.0

Manual Integrations
 APPROVED

Reviewed By :Mahesh Dadoda 07/22/2022
 Supervised By :Semsettin Yesilyurt 07/22/2022




07/22/2022
 Supervised By :Semsettin
 Yesilyurt

07/22/2022



#86
 p-Isopropyltoluene
 Concen: 20.165 ug/l
 RT: 12.012 min Scan# 1792
 Delta R.T. 0.000 min
 Lab File: VX030165.D
 Acq: 21 Jul 2022 16:20


Tgt Ion:119 Resp: 171798
 Ion Ratio Lower Upper
 119 100
 134 26.4 13.9 41.6
 91 24.5 11.8 35.3





#87
1,3-Dichlorobenzene
Concen: 19.407 ug/l
RT: 11.969 min Scan# 1785
Delta R.T. 0.000 min
Lab File: VX030165.D
Acq: 21 Jul 2022 16:20

Instrument :
MSVOA_X
ClientSampleId :





Tgt Ion:146 Resp: 89180
Ion Ratio Lower Upper
146 100
111 41.7 20.3 60.8
148 63.4 31.8 95.3

Manual Integrations
APPROVED

Reviewed By :Mahesh Dadoda 07/22/2022
Supervised By :Semsettin Yesilyurt 07/22/2022




07/22/2022
Supervised By :Semsettin
Yesilyurt


07/22/2022



#88
1,4-Dichlorobenzene
Concen: 18.833 ug/l
RT: 12.043 min Scan# 1797
Delta R.T. 0.001 min
Lab File: VX030165.D
Acq: 21 Jul 2022 16:20


Tgt Ion:146 Resp: 88819
Ion Ratio Lower Upper
146 100
111 41.6 20.3 60.9
148 64.9 32.3 96.8





#89
n-Butylbenzene
Concen: 20.229 ug/l
RT: 12.335 min Scan# 1845
Delta R.T. 0.000 min
Lab File: VX030165.D
Acq: 21 Jul 2022 16:20

Instrument :
MSVOA_X
ClientSampleId :






Tgt Ion: 91 Resp: 154418
Ion Ratio Lower Upper
91 100
92 51.9 26.2 78.5
134 25.4 12.6 37.8

Manual Integrations
APPROVED

Reviewed By :Mahesh Dadoda 07/22/2022
Supervised By :Semsettin Yesilyurt 07/22/2022




07/22/2022
Supervised By :Semsettin
Yesilyurt


07/22/2022



#90
Hexachloroethane
Concen: 18.790 ug/l
RT: 12.542 min Scan# 1879
Delta R.T. 0.000 min
Lab File: VX030165.D
Acq: 21 Jul 2022 16:20


Tgt Ion:117 Resp: 27366
Ion Ratio Lower Upper
117 100
201 65.6 34.2 102.5





#91
1,2-Dichlorobenzene
Concen: 19.419 ug/l
RT: 12.335 min Scan# 1845
Delta R.T. 0.000 min
Lab File: VX030165.D
Acq: 21 Jul 2022 16:20

Instrument :
MSVOA_X
ClientSampleId :






Tgt Ion:146 Resp: 90668
Ion Ratio Lower Upper
146 100
111 41.6 20.7 62.1
148 65.1 32.0 96.2

Manual Integrations
APPROVED

Reviewed By :Mahesh Dadoda 07/22/2022
Supervised By :Semsettin Yesilyurt 07/22/2022




07/22/2022
Supervised By :Semsettin
Yesilyurt


07/22/2022



#92
1,2-Dibromo-3-Chloropropane
Concen: 20.545 ug/l
RT: 12.945 min Scan# 1945
Delta R.T. -0.000 min
Lab File: VX030165.D
Acq: 21 Jul 2022 16:20

Tgt Ion: 75 Resp: 16825
Ion Ratio Lower Upper
75 100
155 78.6 45.0 134.8
157 102.8 56.9 170.7





#93

1,2,4-Trichlorobenzene

Concen: 17.965 ug/l

RT: 13.591 min Scan# 2051

Delta R.T. 0.000 min

Lab File: VX030165.D

Acq: 21 Jul 2022 16:20

Instrument :

MSVOA_X

ClientSampleId :

Tgt Ion:180	Resp:	51520
Ion Ratio	Lower	Upper
180	100	
182	94.2	48.2 144.6
145	32.2	15.9 47.6

Manual Integrations





APPROVED

Reviewed By :Mahesh Dadoda 07/22/2022
Supervised By :Semsettin Yesilyurt 07/22/2022

07/22/2022

Supervised By :Semsettin Yesilyurt

07/22/2022



#94

Hexachlorobutadiene

Concen: 18.456 ug/l




RT: 13.725 min Scan# 2073


Delta R.T. 0.000 min

Lab File: VX030165.D

Acq: 21 Jul 2022 16:20


Tgt Ion:225	Resp:	21131
Ion Ratio	Lower	Upper
225	100	
223	63.0	30.6 91.6
227	62.1	31.9 95.9





#95
Naphthalene
Concen: 18.609 ug/l
RT: 13.780 min Scan# 2082
Delta R.T. 0.000 min
Lab File: VX030165.D
Acq: 21 Jul 2022 16:20

Instrument :
MSVOA_X
ClientSampleId :






Tgt Ion:128 Resp: 211824
Ion Ratio Lower Upper
128 100
127 12.3 10.2 15.2
129 10.6 8.6 13.0

Manual Integrations
APPROVED

Reviewed By :Mahesh Dadoda 07/22/2022
Supervised By :Semsettin Yesilyurt 07/22/2022

07/22/2022
Supervised By :Semsettin
Yesilyurt

07/22/2022



#96
1,2,3-Trichlorobenzene
Concen: 18.344 ug/l
RT: 13.963 min Scan# 2112
Delta R.T. -0.000 min
Lab File: VX030165.D
Acq: 21 Jul 2022 16:20

Tgt Ion:180 Resp: 53528
Ion Ratio Lower Upper
180 100
182 95.0 47.0 141.0
145 34.3 17.0 50.9

