



Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX072122\
 Data File : VX030129.D
 Acq On : 20 Jul 2022 13:04
 Operator : JC/MD
 Sample : BFB
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X071222W.M
 Title : SW846 8260
 Last Update : Wed Jul 13 09:05:23 2022



AutoFind: Scans 1639, 1640, 1641; Background Corrected with Scan 1631

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	18.3	10235	PASS
75	95	30	60	48.3	27003	PASS
95	95	100	100	100.0	55947	PASS
96	95	5	9	6.4	3576	PASS
173	174	0.00	2	1.0	444	PASS
174	95	50	100	76.4	42744	PASS
175	174	5	9	7.3	3120	PASS
176	174	95	101	96.1	41061	PASS
177	176	5	9	6.2	2551	PASS