



Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX072123\
Data File : VX036599.D
Acq On : 21 Jul 2023 08:35
Operator : JC/MD
Sample : BFB
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X072123W.M
Title : SW846 8260
Last Update : Fri Jul 21 12:14:12 2023



AutoFind: Scans 1638, 1639, 1640; Background Corrected with Scan 1627

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	19.2	10627	PASS
75	95	30	60	50.5	27912	PASS
95	95	100	100	100.0	55243	PASS
96	95	5	9	6.6	3652	PASS
173	174	0.00	2	0.9	344	PASS
174	95	50	100	72.2	39888	PASS
175	174	5	9	7.3	2911	PASS
176	174	95	101	96.2	38376	PASS
177	176	5	9	6.4	2461	PASS