

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX072219\
 Data File : VX011020.D
 Acq On : 22 Jul 2019 11:07
 Operator : JC/SP
 Sample : VX0722WBL01
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0722WBL01

Quant Time: Jul 23 04:54:16 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X071619W.M
 Quant Title : SW846 8260
 QLast Update : Sat Jul 20 04:05:59 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.66	168	216913	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.85	114	341801	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	310418	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	138212	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.06	65	118261	51.25	ug/l	0.00
Spiked Amount	50.000		Recovery	=	102.50%	
35) Dibromofluoromethane	5.49	113	109390	50.42	ug/l	0.00
Spiked Amount	50.000		Recovery	=	100.84%	
50) Toluene-d8	8.71	98	407918	50.13	ug/l	0.00
Spiked Amount	50.000		Recovery	=	100.26%	
62) 4-Bromofluorobenzene	11.13	95	141074	47.17	ug/l	0.00
Spiked Amount	50.000		Recovery	=	94.34%	

Target Compounds	Qvalue
------------------	--------

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX072219\
Data File : VX011020.D
Acq On : 22 Jul 2019 11:07
Operator : JC/SP
Sample : VX0722WBL01
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX0722WBL01

Quant Time: Jul 23 04:54:16 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X071619W.M
Quant Title : SW846 8260
QLast Update : Sat Jul 20 04:05:59 2019
Response via : Initial Calibration

