

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX072219\
 Data File : VX011024.D
 Acq On : 22 Jul 2019 12:42
 Operator : JC/SP
 Sample : K3615-05ME
 Misc : 5.06g/5mL/100uL/5.00mL/MSVOA_X/MEOH
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 HR-06-071719-CME

Quant Time: Jul 23 06:34:40 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X071619W.M
 Quant Title : SW846 8260
 QLast Update : Sat Jul 20 04:05:59 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.65	168	235930	50.00	ug/l	-0.01
34) 1,4-Difluorobenzene	6.85	114	365872	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	323840	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	184053	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.06	65	126109	50.24	ug/l	0.00
Spiked Amount	50.000		Recovery	=	100.48%	
35) Dibromofluoromethane	5.50	113	109162	47.01	ug/l	0.00
Spiked Amount	50.000		Recovery	=	94.02%	
50) Toluene-d8	8.71	98	428604	49.21	ug/l	0.00
Spiked Amount	50.000		Recovery	=	98.42%	
62) 4-Bromofluorobenzene	11.14	95	193070	60.31	ug/l	0.00
Spiked Amount	50.000		Recovery	=	120.62%	
Target Compounds						
52) Toluene	8.78	92	7353	1.216	ug/l	99
67) Ethyl Benzene	10.25	91	117406	10.370	ug/l	99
68) m/p-Xylenes	10.36	106	232539	53.546	ug/l	99
69) o-Xylene	10.70	106	202618	47.805	ug/l	98
73) Isopropylbenzene	11.02	105	114875	9.237	ug/l	100
78) n-propylbenzene	11.36	91	738446	53.739	ug/l	99
80) 1,3,5-Trimethylbenzene	11.51	105	2043109	196.606	ug/l	99
83) tert-Butylbenzene	11.77	119	22527m	2.207	ug/l	
84) 1,2,4-Trimethylbenzene	11.81	105	4673618	446.212	ug/l	99
85) sec-Butylbenzene	11.94	105	378744	31.666	ug/l	98
86) p-Isopropyltoluene	12.07	119	360934	32.642	ug/l	100
89) n-Butylbenzene	12.39	91	1565233	165.460	ug/l #	51
95) Naphthalene	13.83	128	267621	22.666	ug/l #	88

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX072219\
Data File : VX011024.D
Acq On : 22 Jul 2019 12:42
Operator : JC/SP
Sample : K3615-05ME
Misc : 5.06g/5mL/100uL/5.00mL/MSVOA_X/MEOH
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
HR-06-071719-CME

Quant Time: Jul 23 06:34:40 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X071619W.M
Quant Title : SW846 8260
QLast Update : Sat Jul 20 04:05:59 2019
Response via : Initial Calibration

