

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX072221\
 Data File : VX023373.D
 Acq On : 22 Jul 2021 11:55
 Operator : JC/MD
 Sample : IBLK
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 IBLK

Quant Time: Jul 23 06:06:39 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X072221W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jul 23 06:05:38 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	183006	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	304218	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	260658	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	108169	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	113465	49.703	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	99.400%
35) Dibromofluoromethane	5.397	113	94545	49.705	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	99.420%
50) Toluene-d8	8.653	98	354772	50.766	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	101.540%
62) 4-Bromofluorobenzene	11.085	95	117251	49.341	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	98.680%

Target Compounds				Qvalue		
2) Dichlorodifluoromethane	1.167	85	1014	0.602	ug/l	93
3) Chloromethane	1.295	50	1164	0.593	ug/l #	77
4) Vinyl Chloride	1.374	62	888	0.447	ug/l	95
5) Bromomethane	1.618	94	2330	1.977	ug/l	90
6) Chloroethane	1.673	64	295	0.242	ug/l #	58
7) Trichlorofluoromethane	1.886	101	1608	0.503	ug/l #	78
9) 1,1,2-Trichlorotrifluo...	2.325	101	1292	0.713	ug/l #	86
10) Methyl Iodide	2.465	142	2414	1.063	ug/l	91
11) Tert butyl alcohol	2.971	59	563	1.080	ug/l #	73
12) 1,1-Dichloroethene	2.325	96	940	0.523	ug/l	86
13) Acrolein	2.246	56	477	3.153	ug/l #	11
14) Allyl chloride	2.666	41	1407	0.430	ug/l #	71
15) Acrylonitrile	3.081	53	2025	1.720	ug/l	95
16) Acetone	2.392	43	1981	1.769	ug/l	90
17) Carbon Disulfide	2.514	76	8928	2.014	ug/l #	94
18) Methyl Acetate	2.715	43	969	0.364	ug/l #	63
20) Methylene Chloride	2.794	84	2256	1.036	ug/l #	80
21) trans-1,2-Dichloroethene	3.099	96	1967	1.030	ug/l #	82
23) Vinyl Acetate	3.739	43	3164	0.555	ug/l #	89
25) 2-Butanone	4.574	43	558	0.321	ug/l #	77
27) cis-1,2-Dichloroethene	4.501	96	1200	0.541	ug/l	73
28) Bromochloromethane	4.916	49	539	0.326	ug/l #	91
29) Tetrahydrofuran	5.038	42	845	0.742	ug/l #	45
31) Cyclohexane	5.477	56	1598	0.473	ug/l #	81
36) 1,1-Dichloropropene	5.696	75	1813	0.659	ug/l	95
38) Carbon Tetrachloride	5.690	117	767	0.274	ug/l #	69
39) Methylcyclohexane	7.385	83	1962	0.565	ug/l	96
40) Benzene	6.050	78	1767	0.217	ug/l	95
42) 1,2-Dichloroethane	6.092	62	696	0.231	ug/l #	73
44) Trichloroethene	7.141	130	1683	0.751	ug/l	66
46) Dibromomethane	7.598	93	583	0.398	ug/l #	81
51) 4-Methyl-2-Pentanone	8.586	43	853	0.259	ug/l #	70
52) Toluene	8.726	92	1508	0.298	ug/l	95
53) t-1,3-Dichloropropene	8.982	75	1121	2.140	ug/l #	59
54) cis-1,3-Dichloropropene	8.379	75	754	0.229	ug/l #	88
58) 2-Chloroethyl Vinyl ether	8.245	63	914	0.526	ug/l	99

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX072221\
 Data File : VX023373.D
 Acq On : 22 Jul 2021 11:55
 Operator : JC/MD
 Sample : IBLK
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 IBLK

Quant Time: Jul 23 06:06:39 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X072221W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jul 23 06:05:38 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


59) 2-Hexanone	9.439	43	1004	0.402	ug/l	80
60) Dibromochloromethane	9.531	129	140	2.042	ug/l	90
61) 1,2-Dibromoethane	9.616	107	571	0.261	ug/l	93
64) Tetrachloroethene	9.275	164	1463	0.705	ug/l	91
65) Chlorobenzene	10.086	112	2271	0.426	ug/l #	87
66) 1,1,1,2-Tetrachloroethane	10.171	131	137	1.040	ug/l #	49
67) Ethyl Benzene	10.195	91	3550	0.384	ug/l	95
68) m/p-Xylenes	10.305	106	2829	0.795	ug/l	93
69) o-Xylene	10.647	106	1047	0.297	ug/l	89
70) Styrene	10.665	104	1865	0.334	ug/l #	88
71) Bromoform	10.805	173	71	3.561	ug/l #	29
73) Isopropylbenzene	10.970	105	3495	0.428	ug/l	100
76) 1,2,3-Trichloropropane	11.085	75	59577	25.029	ug/l #	38
77) Bromobenzene	11.201	156	1021	0.519	ug/l	86
78) n-propylbenzene	11.305	91	5813	0.629	ug/l	97
79) 2-Chlorotoluene	11.372	91	2868	0.520	ug/l	98
80) 1,3,5-Trimethylbenzene	11.457	105	2933	0.441	ug/l	86
81) trans-1,4-Dichloro-2-b...	11.085	75	59577	67.299	ug/l #	12
82) 4-Chlorotoluene	11.457	91	4129	0.661	ug/l	97
83) tert-Butylbenzene	11.720	119	3224	0.494	ug/l	87
84) 1,2,4-Trimethylbenzene	11.756	105	2976	0.450	ug/l	94
85) sec-Butylbenzene	11.896	105	4853	0.574	ug/l	99
86) p-Isopropyltoluene	12.012	119	4233	0.620	ug/l	87
87) 1,3-Dichlorobenzene	11.976	146	2940	0.807	ug/l	98
88) 1,4-Dichlorobenzene	11.976	146	2940	0.795	ug/l	96
89) n-Butylbenzene	12.335	91	4579	0.741	ug/l	98
90) Hexachloroethane	12.536	117	284	2.122	ug/l	74
91) 1,2-Dichlorobenzene	12.341	146	2072	0.587	ug/l	99
92) 1,2-Dibromo-3-Chloropr...	12.957	75	93	2.257	ug/l	76
93) 1,2,4-Trichlorobenzene	13.591	180	2523	1.139	ug/l	96
94) Hexachlorobutadiene	13.731	225	1046	1.052	ug/l	88
95) Naphthalene	13.780	128	5565	0.730	ug/l	98
96) 1,2,3-Trichlorobenzene	13.963	180	2277	1.018	ug/l	95


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX072221\
 Data File : VX023373.D
 Acq On : 22 Jul 2021 11:55
 Operator : JC/MD
 Sample : IBLK
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 IBLK

Quant Time: Jul 23 06:06:39 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X072221W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jul 23 06:05:38 2021
 Response via : Initial Calibration









#1
 Pentafluorobenzene
 Concen: 50.000 ug/l
 RT: 5.562 min Scan# 734
 Delta R.T. 0.000 min
 Lab File: VX023373.D
 Acq: 22 Jul 2021 11:55




Instrument :
 MSVOA_X
 ClientSampleId :
 IBLK


Tgt Ion:168 Resp: 183006
 Ion Ratio Lower Upper
 168 100
 99 57.8 45.8 68.6



#2
 Dichlorodifluoromethane
 Concen: 0.602 ug/l
 RT: 1.167 min Scan# 13
 Delta R.T. 0.001 min
 Lab File: VX023373.D
 Acq: 22 Jul 2021 11:55

Tgt Ion: 85 Resp: 1014
 Ion Ratio Lower Upper
 85 100
 87 29.0 16.4 49.2









#3
Chloromethane
Concen: 0.593 ug/l
RT: 1.295 min Scan# 34
Delta R.T. 0.001 min
Lab File: VX023373.D
Acq: 22 Jul 2021 11:55




Instrument :
MSVOA_X
ClientSampled :
IBLK


Tgt Ion: 50 Resp: 1164
Ion Ratio Lower Upper
50 100
52 20.2 26.4 39.6#



#4
Vinyl Chloride
Concen: 0.447 ug/l
RT: 1.374 min Scan# 47
Delta R.T. -0.000 min
Lab File: VX023373.D
Acq: 22 Jul 2021 11:55

Tgt Ion: 62 Resp: 888
Ion Ratio Lower Upper
62 100
64 29.2 25.7 38.5









#5
Bromomethane
Concen: 1.977 ug/l
RT: 1.618 min Scan# 87
Delta R.T. 0.019 min
Lab File: VX023373.D
Acq: 22 Jul 2021 11:55




Instrument :
MSVOA_X
ClientSampled :
IBLK


Tgt Ion: 94 Resp: 2330
Ion Ratio Lower Upper
94 100
96 86.2 76.4 114.6



#6
Chloroethane
Concen: 0.242 ug/l
RT: 1.673 min Scan# 96
Delta R.T. -0.005 min
Lab File: VX023373.D
Acq: 22 Jul 2021 11:55

Tgt Ion: 64 Resp: 295
Ion Ratio Lower Upper
64 100
66 55.2 25.4 38.0#









#7
Trichlorofluoromethane
Concen: 0.503 ug/l
RT: 1.886 min Scan# 131
Delta R.T. -0.000 min
Lab File: VX023373.D
Acq: 22 Jul 2021 11:55




Instrument :
MSVOA_X
ClientSampled :
IBLK


Tgt Ion:101 Resp: 1608
Ion Ratio Lower Upper
101 100
103 49.2 53.2 79.8#



#9
1,1,2-Trichlorotrifluoroethane
Concen: 0.713 ug/l
RT: 2.325 min Scan# 203
Delta R.T. -0.006 min
Lab File: VX023373.D
Acq: 22 Jul 2021 11:55


Tgt Ion:101 Resp: 1292
Ion Ratio Lower Upper
101 100
85 40.4 34.9 52.3
151 93.0 61.4 92.2#








#10
Methyl Iodide
Concen: 1.063 ug/l
RT: 2.465 min Scan# 226
Delta R.T. 0.012 min
Lab File: VX023373.D
Acq: 22 Jul 2021 11:55

Instrument :
MSVOA_X
ClientSampled :
IBLK







Tgt Ion:142 Resp: 2414
Ion Ratio Lower Upper
142 100
127 48.8 33.3 49.9
141 13.8 11.4 17.0



#11
Tert butyl alcohol
Concen: 1.080 ug/l
RT: 2.971 min Scan# 309
Delta R.T. -0.012 min
Lab File: VX023373.D
Acq: 22 Jul 2021 11:55

Tgt Ion: 59 Resp: 563
Ion Ratio Lower Upper
59 100
57 0.0 7.8 11.8#





#12

1,1-Dichloroethene

Concen: 0.523 ug/l

RT: 2.325 min Scan# 203

Delta R.T. 0.006 min

Lab File: VX023373.D

Acq: 22 Jul 2021 11:55





Tgt Ion: 96 Resp: 940

Ion Ratio Lower Upper

96 100

61 133.4 124.6 187.0

98 67.5 50.6 76.0



#13

Acrolein

Concen: 3.153 ug/l

RT: 2.246 min Scan# 190

Delta R.T. 0.007 min

Lab File: VX023373.D




Acq: 22 Jul 2021 11:55


Tgt Ion: 56 Resp: 477

Ion Ratio Lower Upper

56 100

55 143.8 56.4 84.6#









#14
 Allyl chloride
 Concen: 0.430 ug/l
 RT: 2.666 min Scan# 259
 Delta R.T. 0.000 min
 Lab File: VX023373.D
 Acq: 22 Jul 2021 11:55




Instrument :
 MSVOA_X
 ClientSampleId :
 IBLK


Tgt Ion:	41	Resp:	1407
Ion	Ratio	Lower	Upper
41	100		
39	57.1	51.6	77.4
76	0.0	30.5	45.7#



#15
 Acrylonitrile
 Concen: 1.720 ug/l
 RT: 3.081 min Scan# 327
 Delta R.T. 0.006 min
 Lab File: VX023373.D
 Acq: 22 Jul 2021 11:55

Tgt Ion:	53	Resp:	2025
Ion	Ratio	Lower	Upper
53	100		
52	75.8	65.3	97.9
51	36.5	28.2	42.2









#16
Acetone
Concen: 1.769 ug/l
RT: 2.392 min Scan# 214
Delta R.T. 0.006 min
Lab File: VX023373.D
Acq: 22 Jul 2021 11:55




Instrument :
MSVOA_X
ClientSampled :
IBLK


Tgt Ion: 43 Resp: 1981
Ion Ratio Lower Upper
43 100
58 29.3 27.8 41.8



#17
Carbon Disulfide
Concen: 2.014 ug/l
RT: 2.514 min Scan# 234
Delta R.T. -0.000 min
Lab File: VX023373.D
Acq: 22 Jul 2021 11:55

Tgt Ion: 76 Resp: 8928
Ion Ratio Lower Upper
76 100
78 11.2 7.2 10.8#





#18

Methyl Acetate

Concen: 0.364 ug/l

RT: 2.715 min Scan# 267

Delta R.T. 0.006 min

Lab File: VX023373.D

Acq: 22 Jul 2021 11:55

Instrument :

MSVOA_X

ClientSampled :





IBLK

Tgt Ion: 43 Resp: 969

Ion Ratio Lower Upper

43 100

74 8.3 21.9 32.9#



#20

Methylene Chloride

Concen: 1.036 ug/l

RT: 2.794 min Scan# 280

Delta R.T. 0.000 min

Lab File: VX023373.D

Acq: 22 Jul 2021 11:55

Tgt Ion: 84 Resp: 2256




Ion Ratio Lower Upper


84 100

49 142.4 88.6 132.8#

51 40.8 27.0 40.6#

86 67.3 50.3 75.5









#21
trans-1,2-Dichloroethene
Concen: 1.030 ug/l
RT: 3.099 min Scan# 330
Delta R.T. 0.006 min
Lab File: VX023373.D
Acq: 22 Jul 2021 11:55




Instrument :
MSVOA_X
ClientSampleId :
IBLK


Tgt Ion: 96 Resp: 1967
Ion Ratio Lower Upper
96 100
61 106.9 107.7 161.5#
98 58.4 50.4 75.6



#23
Vinyl Acetate
Concen: 0.555 ug/l
RT: 3.739 min Scan# 435
Delta R.T. 0.006 min
Lab File: VX023373.D
Acq: 22 Jul 2021 11:55

Tgt Ion: 43 Resp: 3164
Ion Ratio Lower Upper
43 100
86 8.0 9.9 14.9#





#25

2-Butanone

Concen: 0.321 ug/l

RT: 4.574 min Scan# 572

Delta R.T. 0.006 min

Lab File: VX023373.D

Acq: 22 Jul 2021 11:55

Tgt Ion: 43 Resp: 558

Ion Ratio Lower Upper

43 100





72 17.2 23.7 35.5#

Instrument :

MSVOA_X

ClientSampled :

IBLK



#27

cis-1,2-Dichloroethene

Concen: 0.541 ug/l

RT: 4.501 min Scan# 560

Delta R.T. 0.006 min

Lab File: VX023373.D

Acq: 22 Jul 2021 11:55




Tgt Ion: 96 Resp: 1200


Ion Ratio Lower Upper

96 100

61 122.6 0.0 285.6

98 25.3 0.0 129.6









#28
Bromochloromethane
Concen: 0.326 ug/l
RT: 4.916 min Scan# 628
Delta R.T. 0.006 min
Lab File: VX023373.D
Acq: 22 Jul 2021 11:55




Instrument :
MSVOA_X
ClientSampled :
IBLK


Tgt Ion: 49 Resp: 539
Ion Ratio Lower Upper
49 100
129 0.0 0.0 4.0
130 77.4 68.2 102.2



#29
Tetrahydrofuran
Concen: 0.742 ug/l
RT: 5.038 min Scan# 648
Delta R.T. 0.025 min
Lab File: VX023373.D
Acq: 22 Jul 2021 11:55

Tgt Ion: 42 Resp: 845
Ion Ratio Lower Upper
42 100
72 12.1 41.7 62.5#
71 12.7 38.4 57.6#









#31
Cyclohexane
Concen: 0.473 ug/l
RT: 5.477 min Scan# 720
Delta R.T. -0.000 min
Lab File: VX023373.D
Acq: 22 Jul 2021 11:55




Instrument :
MSVOA_X
ClientSampled :
IBLK


Tgt Ion: 56 Resp: 1598
Ion Ratio Lower Upper
56 100
69 23.0 25.3 37.9#
84 69.9 71.4 107.0#



#33
1,2-Dichloroethane-d4
Concen: 49.703 ug/l
RT: 5.964 min Scan# 800
Delta R.T. 0.000 min
Lab File: VX023373.D
Acq: 22 Jul 2021 11:55

Tgt Ion: 65 Resp: 113465
Ion Ratio Lower Upper
65 100
67 53.1 0.0 105.4









#34
1,4-Difluorobenzene
Concen: 50.000 ug/l
RT: 6.769 min Scan# 932
Delta R.T. 0.000 min
Lab File: VX023373.D
Acq: 22 Jul 2021 11:55




Instrument :
MSVOA_X
ClientSampleId :
IBLK


Tgt Ion:	114	Resp:	304218
Ion Ratio	Lower	Upper	
114	100		
63	20.2	0.0	40.8
88	16.3	0.0	32.8



#35
Dibromofluoromethane
Concen: 49.705 ug/l
RT: 5.397 min Scan# 707
Delta R.T. 0.000 min
Lab File: VX023373.D
Acq: 22 Jul 2021 11:55

Tgt Ion:	113	Resp:	94545
Ion Ratio	Lower	Upper	
113	100		
111	103.9	81.9	122.9
192	19.7	16.7	25.1





#36

1,1-Dichloropropene

Concen: 0.659 ug/l

RT: 5.696 min Scan# 756





Delta R.T. 0.000 min

Lab File: VX023373.D

Acq: 22 Jul 2021 11:55

Tgt Ion:	75	Resp:	1813
Ion	Ratio	Lower	Upper
75	100		
110	32.4	18.1	54.4
77	29.8	25.3	37.9

Instrument :
MSVOA_X
ClientSampleId :
IBLK



#38

Carbon Tetrachloride

Concen: 0.274 ug/l




RT: 5.690 min Scan# 755


Delta R.T. 0.006 min

Lab File: VX023373.D

Acq: 22 Jul 2021 11:55

Tgt Ion:	117	Resp:	767
Ion	Ratio	Lower	Upper
117	100		
119	57.2	76.2	114.2#
121	33.7	24.8	37.2





#39

Methylcyclohexane

Concen: 0.565 ug/l

RT: 7.385 min Scan# 1033

Delta R.T. -0.000 min

Lab File: VX023373.D

Acq: 22 Jul 2021 11:55

Tgt Ion: 83 Resp: 1962





Ion	Ratio	Lower	Upper
83	100		
55	76.9	58.9	88.3
98	50.9	38.1	57.1

Instrument :

MSVOA_X

ClientSampleId :

IBLK



#40

Benzene

Concen: 0.217 ug/l

RT: 6.050 min Scan# 814




Delta R.T. 0.007 min


Lab File: VX023373.D

Acq: 22 Jul 2021 11:55

Tgt Ion: 78 Resp: 1767

Ion	Ratio	Lower	Upper
78	100		
77	21.1	18.9	28.3





#42

1,2-Dichloroethane

Concen: 0.231 ug/l

RT: 6.092 min Scan# 821

Delta R.T. -0.006 min





Lab File: VX023373.D

Acq: 22 Jul 2021 11:55

Tgt Ion: 62 Resp: 696

Ion Ratio Lower Upper

62 100
98 0.0 0.0 20.2



#44

Trichloroethene

Concen: 0.751 ug/l

RT: 7.141 min Scan# 993

Delta R.T. 0.012 min




Lab File: VX023373.D


Acq: 22 Jul 2021 11:55

Tgt Ion: 130 Resp: 1683

Ion Ratio Lower Upper

130 100
95 65.0 0.0 198.0









#46
Dibromomethane
Concen: 0.398 ug/l
RT: 7.598 min Scan# 1068
Delta R.T. 0.012 min
Lab File: VX023373.D
Acq: 22 Jul 2021 11:55




Instrument :
MSVOA_X
ClientSampleId :
IBLK


Tgt Ion: 93 Resp: 583
Ion Ratio Lower Upper
93 100
95 93.5 66.8 100.2
174 78.2 83.1 124.7#



#50
Toluene-d8
Concen: 50.766 ug/l
RT: 8.653 min Scan# 1241
Delta R.T. -0.000 min
Lab File: VX023373.D
Acq: 22 Jul 2021 11:55

Tgt Ion: 98 Resp: 354772
Ion Ratio Lower Upper
98 100
100 63.8 51.9 77.9









#51
4-Methyl-2-Pentanone
Concen: 0.259 ug/l
RT: 8.586 min Scan# 1230
Delta R.T. 0.006 min
Lab File: VX023373.D
Acq: 22 Jul 2021 11:55




Instrument :
MSVOA_X
ClientSampled :
IBLK


Tgt Ion: 43 Resp: 853
Ion Ratio Lower Upper
43 100
58 24.7 35.4 53.0#



#52
Toluene
Concen: 0.298 ug/l
RT: 8.726 min Scan# 1253
Delta R.T. 0.006 min
Lab File: VX023373.D
Acq: 22 Jul 2021 11:55

Tgt Ion: 92 Resp: 1508
Ion Ratio Lower Upper
92 100
91 178.1 137.1 205.7









#53
 t-1,3-Dichloropropene
 Concen: 2.140 ug/l
 RT: 8.982 min Scan# 1295
 Delta R.T. 0.000 min
 Lab File: VX023373.D
 Acq: 22 Jul 2021 11:55




Instrument :
 MSVOA_X
 ClientSampled :
 IBLK


Tgt Ion: 75 Resp: 1121
 Ion Ratio Lower Upper
 75 100
 77 53.6 24.9 37.3#



#54
 cis-1,3-Dichloropropene
 Concen: 0.229 ug/l
 RT: 8.379 min Scan# 1196
 Delta R.T. 0.007 min
 Lab File: VX023373.D
 Acq: 22 Jul 2021 11:55

Tgt Ion: 75 Resp: 754
 Ion Ratio Lower Upper
 75 100
 77 34.1 25.7 38.5
 39 55.4 35.1 52.7#









#58
2-Chloroethyl Vinyl ether
Concen: 0.526 ug/l
RT: 8.245 min Scan# 1174
Delta R.T. -0.005 min
Lab File: VX023373.D
Acq: 22 Jul 2021 11:55




Instrument :
MSVOA_X
ClientSampleId :
IBLK


Tgt Ion: 63 Resp: 914
Ion Ratio Lower Upper
63 100
106 29.1 23.5 35.3



#59
2-Hexanone
Concen: 0.402 ug/l
RT: 9.439 min Scan# 1370
Delta R.T. 0.006 min
Lab File: VX023373.D
Acq: 22 Jul 2021 11:55

Tgt Ion: 43 Resp: 1004
Ion Ratio Lower Upper
43 100
58 46.2 30.9 92.8





#60

Dibromochloromethane

Concen: 2.042 ug/l

RT: 9.531 min Scan# 1385

Delta R.T. 0.006 min

Lab File: VX023373.D





Acq: 22 Jul 2021 11:55

Tgt Ion:129 Resp: 140

Ion Ratio Lower Upper

129 100

127 85.7 38.4 115.2



#61

1,2-Dibromoethane

Concen: 0.261 ug/l

RT: 9.616 min Scan# 1399

Delta R.T. 0.000 min

Lab File: VX023373.D




Acq: 22 Jul 2021 11:55


Tgt Ion:107 Resp: 571

Ion Ratio Lower Upper

107 100

109 101.8 75.8 113.6









#62
4-Bromofluorobenzene
Concen: 49.341 ug/l
RT: 11.085 min Scan# 1640
Delta R.T. 0.000 min
Lab File: VX023373.D
Acq: 22 Jul 2021 11:55




Instrument :
MSVOA_X
ClientSampleId :
IBLK


Tgt Ion: 95 Resp: 117251
Ion Ratio Lower Upper
95 100
174 81.2 0.0 154.6
176 78.8 0.0 155.8



#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 10.055 min Scan# 1471
Delta R.T. 0.000 min
Lab File: VX023373.D
Acq: 22 Jul 2021 11:55


Tgt Ion: 117 Resp: 260658
Ion Ratio Lower Upper
117 100
82 57.0 47.4 71.2
119 30.8 26.6 39.8








#64
Tetrachloroethene
Concen: 0.705 ug/l
RT: 9.275 min Scan# 1343
Delta R.T. -0.000 min
Lab File: VX023373.D
Acq: 22 Jul 2021 11:55

Instrument :
MSVOA_X
ClientSampleId :
IBLK







Tgt Ion:	164	Resp:	1463
Ion	Ratio	Lower	Upper
164	100		
166	121.6	103.4	155.2
129	72.6	70.8	106.2
131	86.7	72.5	108.7



#65
Chlorobenzene
Concen: 0.426 ug/l
RT: 10.086 min Scan# 1476
Delta R.T. 0.001 min
Lab File: VX023373.D
Acq: 22 Jul 2021 11:55

Tgt Ion:	112	Resp:	2271
Ion	Ratio	Lower	Upper
112	100		
114	39.4	25.8	38.6#









#66
1,1,1,2-Tetrachloroethane
Concen: 1.040 ug/l
RT: 10.171 min Scan# 1490
Delta R.T. 0.006 min
Lab File: VX023373.D
Acq: 22 Jul 2021 11:55




Instrument :
MSVOA_X
ClientSampled :
IBLK


Tgt Ion:131 Resp: 137
Ion Ratio Lower Upper
131 100
133 124.1 47.7 143.1
119 0.0 32.1 96.5#



#67
Ethyl Benzene
Concen: 0.384 ug/l
RT: 10.195 min Scan# 1494
Delta R.T. 0.000 min
Lab File: VX023373.D
Acq: 22 Jul 2021 11:55

Tgt Ion: 91 Resp: 3550
Ion Ratio Lower Upper
91 100
106 33.5 24.6 37.0









#68
m/p-Xylenes
Concen: 0.795 ug/l
RT: 10.305 min Scan# 1512
Delta R.T. 0.000 min
Lab File: VX023373.D
Acq: 22 Jul 2021 11:55




Instrument :
MSVOA_X
ClientSampleId :
IBLK


Tgt Ion:106 Resp: 2829
Ion Ratio Lower Upper
106 100
91 192.8 162.8 244.2



#69
o-Xylene
Concen: 0.297 ug/l
RT: 10.647 min Scan# 1568
Delta R.T. 0.001 min
Lab File: VX023373.D
Acq: 22 Jul 2021 11:55

Tgt Ion:106 Resp: 1047
Ion Ratio Lower Upper
106 100
91 198.9 108.1 324.2









#70
Styrene
Concen: 0.334 ug/l
RT: 10.665 min Scan# 1571
Delta R.T. 0.006 min
Lab File: VX023373.D
Acq: 22 Jul 2021 11:55




Instrument :
MSVOA_X
ClientSampled :
IBLK


Tgt Ion:104 Resp: 1865
Ion Ratio Lower Upper
104 100
78 53.7 40.0 60.0
103 66.9 43.4 65.0#



#71
Bromoform
Concen: 3.561 ug/l
RT: 10.805 min Scan# 1594
Delta R.T. 0.000 min
Lab File: VX023373.D
Acq: 22 Jul 2021 11:55

Tgt Ion:173 Resp: 71
Ion Ratio Lower Upper
173 100
175 0.0 23.9 71.8#
254 0.0 0.1 0.1#









#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 12.024 min Scan# 1794
Delta R.T. 0.000 min
Lab File: VX023373.D
Acq: 22 Jul 2021 11:55




Instrument :
MSVOA_X
ClientSampled :
IBLK


Tgt Ion:152 Resp: 108169
Ion Ratio Lower Upper
152 100
115 58.8 41.1 123.3
150 157.1 0.0 349.0



#73
Isopropylbenzene
Concen: 0.428 ug/l
RT: 10.970 min Scan# 1621
Delta R.T. 0.007 min
Lab File: VX023373.D
Acq: 22 Jul 2021 11:55

Tgt Ion:105 Resp: 3495
Ion Ratio Lower Upper
105 100
120 26.5 13.2 39.6









#76
1,2,3-Trichloropropane
Concen: 25.029 ug/l
RT: 11.085 min Scan# 1640
Delta R.T. -0.159 min
Lab File: VX023373.D
Acq: 22 Jul 2021 11:55




Instrument :
MSVOA_X
ClientSampled :
IBLK


Tgt Ion: 75 Resp: 59577
Ion Ratio Lower Upper
75 100
77 1.0 19.1 57.5#



#77
Bromobenzene
Concen: 0.519 ug/l
RT: 11.201 min Scan# 1659
Delta R.T. 0.000 min
Lab File: VX023373.D
Acq: 22 Jul 2021 11:55

Tgt Ion: 156 Resp: 1021
Ion Ratio Lower Upper
156 100
77 141.7 82.8 248.6
158 92.1 49.1 147.4









#78
n-propylbenzene
Concen: 0.629 ug/l
RT: 11.305 min Scan# 1676
Delta R.T. -0.006 min
Lab File: VX023373.D
Acq: 22 Jul 2021 11:55




Instrument :
MSVOA_X
ClientSampled :
IBLK


Tgt Ion: 91 Resp: 5813
Ion Ratio Lower Upper
91 100
120 21.7 11.5 34.5



#79
2-Chlorotoluene
Concen: 0.520 ug/l
RT: 11.372 min Scan# 1687
Delta R.T. 0.006 min
Lab File: VX023373.D
Acq: 22 Jul 2021 11:55

Tgt Ion: 91 Resp: 2868
Ion Ratio Lower Upper
91 100
126 32.1 16.6 49.8









#80
1,3,5-Trimethylbenzene
Concen: 0.441 ug/l
RT: 11.457 min Scan# 1701
Delta R.T. 0.000 min
Lab File: VX023373.D
Acq: 22 Jul 2021 11:55




Instrument :
MSVOA_X
ClientSampled :
IBLK


Tgt Ion:105 Resp: 2933
Ion Ratio Lower Upper
105 100
120 56.9 23.8 71.5



#81
trans-1,4-Dichloro-2-butene
Concen: 67.299 ug/l
RT: 11.085 min Scan# 1640
Delta R.T. 0.061 min
Lab File: VX023373.D
Acq: 22 Jul 2021 11:55

Tgt Ion: 75 Resp: 59577
Ion Ratio Lower Upper
75 100
53 0.0 74.5 111.7#
89 0.0 37.9 56.9#









#82
4-Chlorotoluene
Concen: 0.661 ug/l
RT: 11.457 min Scan# 1701
Delta R.T. 0.000 min
Lab File: VX023373.D
Acq: 22 Jul 2021 11:55




Instrument :
MSVOA_X
ClientSampled :
IBLK


Tgt Ion: 91 Resp: 4129
Ion Ratio Lower Upper
91 100
126 31.4 14.9 44.5



#83
tert-Butylbenzene
Concen: 0.494 ug/l
RT: 11.720 min Scan# 1744
Delta R.T. 0.001 min
Lab File: VX023373.D
Acq: 22 Jul 2021 11:55

Tgt Ion: 119 Resp: 3224
Ion Ratio Lower Upper
119 100
91 67.2 27.9 83.7
134 20.0 11.3 33.9









#84
1,2,4-Trimethylbenzene
Concen: 0.450 ug/l
RT: 11.756 min Scan# 1750
Delta R.T. 0.000 min
Lab File: VX023373.D
Acq: 22 Jul 2021 11:55




Instrument :
MSVOA_X
ClientSampleId :
IBLK


Tgt Ion:105 Resp: 2976
Ion Ratio Lower Upper
105 100
120 49.0 22.4 67.3



#85
sec-Butylbenzene
Concen: 0.574 ug/l
RT: 11.896 min Scan# 1773
Delta R.T. 0.000 min
Lab File: VX023373.D
Acq: 22 Jul 2021 11:55

Tgt Ion:105 Resp: 4853
Ion Ratio Lower Upper
105 100
134 20.4 10.0 29.8









#86
p-Isopropyltoluene
Concen: 0.620 ug/l
RT: 12.012 min Scan# 1792
Delta R.T. 0.000 min
Lab File: VX023373.D
Acq: 22 Jul 2021 11:55




Instrument :
MSVOA_X
ClientSampleId :
IBLK


Tgt Ion:	119	Resp:	4233
Ion	Ratio	Lower	Upper
119	100		
134	29.3	12.9	38.7
91	33.1	11.7	35.0



#87
1,3-Dichlorobenzene
Concen: 0.807 ug/l
RT: 11.976 min Scan# 1786
Delta R.T. 0.001 min
Lab File: VX023373.D
Acq: 22 Jul 2021 11:55

Tgt Ion:	146	Resp:	2940
Ion	Ratio	Lower	Upper
146	100		
111	42.4	20.2	60.5
148	65.3	31.9	95.7









#88
1,4-Dichlorobenzene
Concen: 0.795 ug/l
RT: 11.976 min Scan# 1786
Delta R.T. -0.066 min
Lab File: VX023373.D
Acq: 22 Jul 2021 11:55




Instrument :
MSVOA_X
ClientSampled :
IBLK


Tgt Ion:146 Resp: 2940
Ion Ratio Lower Upper
146 100
111 42.4 19.6 58.8
148 65.3 31.6 95.0



#89
n-Butylbenzene
Concen: 0.741 ug/l
RT: 12.335 min Scan# 1845
Delta R.T. 0.000 min
Lab File: VX023373.D
Acq: 22 Jul 2021 11:55

Tgt Ion: 91 Resp: 4579
Ion Ratio Lower Upper
91 100
92 56.3 27.4 82.0
134 26.3 13.2 39.5









#90
Hexachloroethane
Concen: 2.122 ug/l
RT: 12.536 min Scan# 1878
Delta R.T. -0.006 min
Lab File: VX023373.D
Acq: 22 Jul 2021 11:55




Instrument :
MSVOA_X
ClientSampleId :
IBLK


Tgt Ion:117 Resp: 284
Ion Ratio Lower Upper
117 100
201 57.7 40.4 121.2



#91
1,2-Dichlorobenzene
Concen: 0.587 ug/l
RT: 12.341 min Scan# 1846
Delta R.T. 0.000 min
Lab File: VX023373.D
Acq: 22 Jul 2021 11:55

Tgt Ion:146 Resp: 2072
Ion Ratio Lower Upper
146 100
111 39.8 20.8 62.2
148 63.3 31.9 95.5









#92
1,2-Dibromo-3-Chloropropane
Concen: 2.257 ug/l
RT: 12.957 min Scan# 1947
Delta R.T. 0.012 min
Lab File: VX023373.D
Acq: 22 Jul 2021 11:55




Instrument :
MSVOA_X
ClientSampled :
IBLK


Tgt Ion: 75 Resp: 93
Ion Ratio Lower Upper
75 100
155 61.3 45.3 135.9
157 95.7 57.7 173.1



#93
1,2,4-Trichlorobenzene
Concen: 1.139 ug/l
RT: 13.591 min Scan# 2051
Delta R.T. 0.000 min
Lab File: VX023373.D
Acq: 22 Jul 2021 11:55

Tgt Ion:180 Resp: 2523
Ion Ratio Lower Upper
180 100
182 91.4 47.7 143.1
145 28.5 15.0 45.0









#94
Hexachlorobutadiene
Concen: 1.052 ug/l
RT: 13.731 min Scan# 2074
Delta R.T. 0.000 min
Lab File: VX023373.D
Acq: 22 Jul 2021 11:55




Instrument :
MSVOA_X
ClientSampleId :
IBLK


Tgt Ion:	225	Resp:	1046
Ion	Ratio	Lower	Upper
225	100		
223	49.2	31.1	93.2
227	58.1	31.6	95.0



#95
Naphthalene
Concen: 0.730 ug/l
RT: 13.780 min Scan# 2082
Delta R.T. 0.000 min
Lab File: VX023373.D
Acq: 22 Jul 2021 11:55

Tgt Ion:	128	Resp:	5565
Ion	Ratio	Lower	Upper
128	100		
127	12.0	10.3	15.5
129	10.2	8.8	13.2





#96
 1,2,3-Trichlorobenzene
 Concen: 1.018 ug/l
 RT: 13.963 min Scan# 2112
 Delta R.T. 0.000 min
 Lab File: VX023373.D
 Acq: 22 Jul 2021 11:55

Instrument :
 MSVOA_X
 ClientSampleId :
 IBLK

Tgt Ion:180 Resp: 2277
 Ion Ratio Lower Upper
 180 100
 182 91.3 47.6 142.9
 145 35.6 16.1 48.3

