

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX072221\
 Data File : VX023375.D
 Acq On : 22 Jul 2021 15:41
 Operator : JC/MD
 Sample : VX0722WBL01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0722WBL01

Quant Time: Jul 23 06:07:22 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X072221W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jul 23 06:05:38 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	182009	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	308683	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	271665	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	107198	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	115889	51.043	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	= 102.080%	
35) Dibromofluoromethane	5.391	113	96979	50.248	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 100.500%	
50) Toluene-d8	8.653	98	364379	51.386	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	= 102.780%	
62) 4-Bromofluorobenzene	11.085	95	118048	48.958	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 97.920%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX072221\
Data File : VX023375.D
Acq On : 22 Jul 2021 15:41
Operator : JC/MD
Sample : VX0722WBL01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX0722WBL01

Quant Time: Jul 23 06:07:22 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X072221W.M
Quant Title : SW846 8260
QLast Update : Fri Jul 23 06:05:38 2021
Response via : Initial Calibration

