

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX072221\
 Data File : VX023381.D
 Acq On : 22 Jul 2021 18:12
 Operator : JC/MD
 Sample : PB137853ZHE#02
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB137853ZHE#02

Quant Time: Jul 23 06:08:48 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X072221W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jul 23 06:05:38 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	173556	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	288863	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	249099	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	102885	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.971	65	110203	50.903	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	101.800%
35) Dibromofluoromethane	5.397	113	91905	50.886	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	101.780%
50) Toluene-d8	8.653	98	339110	51.104	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	102.200%
62) 4-Bromofluorobenzene	11.085	95	107787	47.770	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	95.540%
Target Compounds						
						Qvalue
16) Acetone	2.386	43	7283	6.858	ug/l	97
20) Methylene Chloride	2.794	84	7655	3.706	ug/l	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX072221\
Data File : VX023381.D
Acq On : 22 Jul 2021 18:12
Operator : JC/MD
Sample : PB137853ZHE#02
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB137853ZHE#02

Quant Time: Jul 23 06:08:48 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X072221W.M
Quant Title : SW846 8260
QLast Update : Fri Jul 23 06:05:38 2021
Response via : Initial Calibration

