

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX072221\
 Data File : VX023383.D
 Acq On : 22 Jul 2021 18:59
 Operator : JC/MD
 Sample : M3118-05
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 20210588-SPLP-AMENDED-COM-2

Quant Time: Jul 23 06:09:12 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X072221W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jul 23 06:05:38 2021
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	175481	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	290564	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	256360	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	103261	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	111044	50.729	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	101.460%
35) Dibromofluoromethane	5.397	113	76974	42.369	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	84.740%
50) Toluene-d8	8.652	98	347619	52.080	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	104.160%
62) 4-Bromofluorobenzene	11.085	95	112564	49.595	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	99.180%
Target Compounds						
						Qvalue
16) Acetone	2.391	43	23795	22.160	ug/l	95
20) Methylene Chloride	2.794	84	10070	4.821	ug/l #	79

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX072221\
Data File : VX023383.D
Acq On : 22 Jul 2021 18:59
Operator : JC/MD
Sample : M3118-05
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
20210588-SPLP-AMENDED-COM-2

Quant Time: Jul 23 06:09:12 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X072221W.M
Quant Title : SW846 8260
QLast Update : Fri Jul 23 06:05:38 2021
Response via : Initial Calibration

