

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX072221\
 Data File : VX023384.D
 Acq On : 22 Jul 2021 19:23
 Operator : JC/MD
 Sample : PB137858TB
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB137858TB

Quant Time: Jul 23 06:09:25 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X072221W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jul 23 06:05:38 2021
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	176799	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	298076	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	257128	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	101203	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.970	65	114594	51.960	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	= 103.920%	
35) Dibromofluoromethane	5.397	113	94218	50.554	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 101.100%	
50) Toluene-d8	8.653	98	353215	51.585	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	= 103.160%	
62) 4-Bromofluorobenzene	11.085	95	110656	47.525	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 95.060%	
Target Compounds						Qvalue
16) Acetone	2.386	43	31480	29.098	ug/l	95
20) Methylene Chloride	2.794	84	16186	7.692	ug/l	92
43) Isopropyl Acetate	6.348	43	27237	5.170	ug/l	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX072221\
Data File : VX023384.D
Acq On : 22 Jul 2021 19:23
Operator : JC/MD
Sample : PB137858TB
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB137858TB

Quant Time: Jul 23 06:09:25 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X072221W.M
Quant Title : SW846 8260
QLast Update : Fri Jul 23 06:05:38 2021
Response via : Initial Calibration

