

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX072221\
 Data File : VX023392.D
 Acq On : 22 Jul 2021 23:19
 Operator : JC/MD
 Sample : M3145-10
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 30 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 152140-MW-2S(OFFSIDE)

Quant Time: Jul 23 06:11:17 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X072221W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jul 23 06:05:38 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	166797	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	273724	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.061	117	235879	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	81683	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	102617	49.320	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	98.640%
35) Dibromofluoromethane	5.397	113	85383	49.889	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	99.780%
50) Toluene-d8	8.653	98	327791	52.131	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	104.260%
62) 4-Bromofluorobenzene	11.085	95	98235	45.944	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	91.880%
Target Compounds						
27) cis-1,2-Dichloroethene	4.495	96	6872	3.399	ug/l	Qvalue 89

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX072221\
Data File : VX023392.D
Acq On : 22 Jul 2021 23:19
Operator : JC/MD
Sample : M3145-10
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 30 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
152140-MW-2S(OFFSIDE)

Quant Time: Jul 23 06:11:17 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X072221W.M
Quant Title : SW846 8260
QLast Update : Fri Jul 23 06:05:38 2021
Response via : Initial Calibration

