

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_X\Data\VX072221\  
 Data File : VX023394.D  
 Acq On : 23 Jul 2021 00:06  
 Operator : JC/MD  
 Sample : M3145-12  
 Misc : 5.0mL/MSVOA\_X/WATER  
 ALS Vial : 32 Sample Multiplier: 1

Instrument :  
 MSVOA\_X  
 ClientSampleId :  
 152140-MW-2D(OFFSITE)

Quant Time: Jul 23 06:11:43 2021  
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_X\Method\82X072221W.M  
 Quant Title : SW846 8260  
 QLast Update : Fri Jul 23 06:05:38 2021  
 Response via : Initial Calibration

| Compound                    | R.T.   | QIon  | Response | Conc     | Units      | Dev(Min) |
|-----------------------------|--------|-------|----------|----------|------------|----------|
| -----                       |        |       |          |          |            |          |
| Internal Standards          |        |       |          |          |            |          |
| 1) Pentafluorobenzene       | 5.562  | 168   | 170618   | 50.000   | ug/l       | 0.00     |
| 34) 1,4-Difluorobenzene     | 6.769  | 114   | 289822   | 50.000   | ug/l       | 0.00     |
| 63) Chlorobenzene-d5        | 10.055 | 117   | 250498   | 50.000   | ug/l       | 0.00     |
| 72) 1,4-Dichlorobenzene-d4  | 12.024 | 152   | 100683   | 50.000   | ug/l       | 0.00     |
| System Monitoring Compounds |        |       |          |          |            |          |
| 33) 1,2-Dichloroethane-d4   | 5.964  | 65    | 110735   | 52.030   | ug/l       | 0.00     |
| Spiked Amount               | 50.000 | Range | 78 - 117 | Recovery | = 104.060% |          |
| 35) Dibromofluoromethane    | 5.397  | 113   | 91455    | 50.469   | ug/l       | 0.00     |
| Spiked Amount               | 50.000 | Range | 75 - 124 | Recovery | = 100.940% |          |
| 50) Toluene-d8              | 8.653  | 98    | 340125   | 51.088   | ug/l       | 0.00     |
| Spiked Amount               | 50.000 | Range | 92 - 112 | Recovery | = 102.180% |          |
| 62) 4-Bromofluorobenzene    | 11.085 | 95    | 107208   | 47.356   | ug/l       | 0.00     |
| Spiked Amount               | 50.000 | Range | 83 - 123 | Recovery | = 94.720%  |          |
| Target Compounds            |        |       |          |          |            |          |
|                             |        |       |          |          |            | Qvalue   |
| 16) Acetone                 | 2.392  | 43    | 3315     | 3.175    | ug/l       | 92       |
| 27) cis-1,2-Dichloroethene  | 4.495  | 96    | 5049     | 2.441    | ug/l       | 93       |
| -----                       |        |       |          |          |            |          |

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_X\Data\VX072221\  
Data File : VX023394.D  
Acq On : 23 Jul 2021 00:06  
Operator : JC/MD  
Sample : M3145-12  
Misc : 5.0mL/MSVOA\_X/WATER  
ALS Vial : 32 Sample Multiplier: 1

Instrument :  
MSVOA\_X  
ClientSampleId :  
152140-MW-2D(OFFSITE)

Quant Time: Jul 23 06:11:43 2021  
Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_X\Method\82X072221W.M  
Quant Title : SW846 8260  
QLast Update : Fri Jul 23 06:05:38 2021  
Response via : Initial Calibration

