

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX072221\
 Data File : VX023396.D
 Acq On : 23 Jul 2021 00:53
 Operator : JC/MD
 Sample : M3145-16
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 34 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 152140-MW-F

Quant Time: Jul 23 06:12:07 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X072221W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jul 23 06:05:38 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	175458	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	293521	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	255104	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	101950	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.970	65	113849	52.017	ug/l	0.00
Spiked Amount	50.000	Range 78 - 117	Recovery	= 104.040%		
35) Dibromofluoromethane	5.397	113	92018	50.140	ug/l	0.00
Spiked Amount	50.000	Range 75 - 124	Recovery	= 100.280%		
50) Toluene-d8	8.653	98	347563	51.547	ug/l	0.00
Spiked Amount	50.000	Range 92 - 112	Recovery	= 103.100%		
62) 4-Bromofluorobenzene	11.085	95	108325	47.246	ug/l	0.00
Spiked Amount	50.000	Range 83 - 123	Recovery	= 94.500%		
Target Compounds						
					Qvalue	
16) Acetone	2.386	43	3257	3.034	ug/l	92
20) Methylene Chloride	2.794	84	1335	0.639	ug/l #	81
30) Chloroform	5.105	83	14641	4.206	ug/l	96
47) Bromodichloromethane	7.830	83	2011	0.790	ug/l #	94
60) Dibromochloromethane	9.525	129	623	2.256	ug/l	79
64) Tetrachloroethene	9.281	164	1082	0.533	ug/l #	79

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX072221\
Data File : VX023396.D
Acq On : 23 Jul 2021 00:53
Operator : JC/MD
Sample : M3145-16
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 34 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
152140-MW-F

Quant Time: Jul 23 06:12:07 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X072221W.M
Quant Title : SW846 8260
QLast Update : Fri Jul 23 06:05:38 2021
Response via : Initial Calibration

