

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX072221\
 Data File : VX023397.D
 Acq On : 23 Jul 2021 01:17
 Operator : JC/MD
 Sample : M3145-17
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 35 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 152140-MW-C

Quant Time: Jul 23 06:12:18 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X072221W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jul 23 06:05:38 2021
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	172444	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	289251	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	245639	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	95974	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	110199	51.229	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	= 102.460%	
35) Dibromofluoromethane	5.397	113	90937	50.282	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 100.560%	
50) Toluene-d8	8.653	98	339200	51.049	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	= 102.100%	
62) 4-Bromofluorobenzene	11.085	95	104651	46.318	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 92.640%	
Target Compounds						
					Qvalue	
16) Acetone	2.386	43	2167	2.054	ug/l	100
20) Methylene Chloride	2.794	84	687	0.335	ug/l	95
44) Trichloroethene	7.135	130	1832	0.860	ug/l	73
64) Tetrachloroethene	9.281	164	1327	0.679	ug/l #	75

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX072221\
Data File : VX023397.D
Acq On : 23 Jul 2021 01:17
Operator : JC/MD
Sample : M3145-17
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 35 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
152140-MW-C

Quant Time: Jul 23 06:12:18 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X072221W.M
Quant Title : SW846 8260
QLast Update : Fri Jul 23 06:05:38 2021
Response via : Initial Calibration

