

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX072221\
 Data File : VX023398.D
 Acq On : 23 Jul 2021 01:40
 Operator : JC/MD
 Sample : M3145-18
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 36 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 152140-FD-1

Quant Time: Jul 23 06:12:29 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X072221W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jul 23 06:05:38 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.568	168	166840	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	277824	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	234888	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	95375	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	107839	51.816	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	103.640%
35) Dibromofluoromethane	5.397	113	88089	50.711	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	101.420%
50) Toluene-d8	8.653	98	323519	50.692	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	101.380%
62) 4-Bromofluorobenzene	11.085	95	101619	46.825	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	93.660%
Target Compounds						
						Qvalue
16) Acetone	2.392	43	2629	2.575	ug/l	# 88
20) Methylene Chloride	2.788	84	795	0.400	ug/l	87
44) Trichloroethene	7.135	130	1694	0.828	ug/l	85
64) Tetrachloroethene	9.281	164	1344	0.719	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX072221\
Data File : VX023398.D
Acq On : 23 Jul 2021 01:40
Operator : JC/MD
Sample : M3145-18
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 36 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
152140-FD-1

Quant Time: Jul 23 06:12:29 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X072221W.M
Quant Title : SW846 8260
QLast Update : Fri Jul 23 06:05:38 2021
Response via : Initial Calibration

