

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX072221\
 Data File : VX023399.D
 Acq On : 23 Jul 2021 02:04
 Operator : JC/MD
 Sample : M3145-19
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 37 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 152140-MW-14D

Quant Time: Jul 23 06:12:42 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X072221W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jul 23 06:05:38 2021
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	167191	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	284618	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	246236	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	93358	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	108680	52.111	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	= 104.220%	
35) Dibromofluoromethane	5.397	113	89040	50.035	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 100.060%	
50) Toluene-d8	8.653	98	339627	51.946	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	= 103.900%	
62) 4-Bromofluorobenzene	11.085	95	104302	46.915	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 93.820%	
Target Compounds						Qvalue
16) Acetone	2.386	43	3724	3.640	ug/l	89
44) Trichloroethene	7.141	130	953	0.454	ug/l	93
64) Tetrachloroethene	9.281	164	797	0.407	ug/l #	72

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX072221\
Data File : VX023399.D
Acq On : 23 Jul 2021 02:04
Operator : JC/MD
Sample : M3145-19
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 37 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
152140-MW-14D

Quant Time: Jul 23 06:12:42 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X072221W.M
Quant Title : SW846 8260
QLast Update : Fri Jul 23 06:05:38 2021
Response via : Initial Calibration

