

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX072221\
 Data File : VX023400.D
 Acq On : 23 Jul 2021 02:27
 Operator : JC/MD
 Sample : M3145-21
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 38 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 152140-MW-3S(ONSITE)

Quant Time: Jul 23 06:12:54 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X072221W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jul 23 06:05:38 2021
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	172657	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	292066	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	251503	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	94440	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	110974	51.526	ug/l	0.00
Spiked Amount	50.000	Range 78 - 117	Recovery	= 103.060%		
35) Dibromofluoromethane	5.397	113	91994	50.377	ug/l	0.00
Spiked Amount	50.000	Range 75 - 124	Recovery	= 100.760%		
50) Toluene-d8	8.653	98	347125	51.738	ug/l	0.00
Spiked Amount	50.000	Range 92 - 112	Recovery	= 103.480%		
62) 4-Bromofluorobenzene	11.085	95	104246	45.694	ug/l	0.00
Spiked Amount	50.000	Range 83 - 123	Recovery	= 91.380%		
Target Compounds						
					Qvalue	
16) Acetone	2.392	43	2884	2.730	ug/l #	82
20) Methylene Chloride	2.788	84	766	0.373	ug/l #	82
27) cis-1,2-Dichloroethene	4.501	96	938	0.448	ug/l	65
44) Trichloroethene	7.135	130	1174	0.546	ug/l	78
64) Tetrachloroethene	9.281	164	2653	1.325	ug/l	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX072221\
Data File : VX023400.D
Acq On : 23 Jul 2021 02:27
Operator : JC/MD
Sample : M3145-21
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 38 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
152140-MW-3S(ONSITE)

Quant Time: Jul 23 06:12:54 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X072221W.M
Quant Title : SW846 8260
QLast Update : Fri Jul 23 06:05:38 2021
Response via : Initial Calibration

