

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX072221\
 Data File : VX023401.D
 Acq On : 23 Jul 2021 02:51
 Operator : JC/MD
 Sample : M3145-22
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 39 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 152140-FD-3

Quant Time: Jul 23 06:13:05 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X072221W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jul 23 06:05:38 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	171412	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	287746	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	246349	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	100343	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	110033	51.460	ug/l	0.00
Spiked Amount	50.000	Range 78 - 117	Recovery	= 102.920%		
35) Dibromofluoromethane	5.397	113	89996	50.022	ug/l	0.00
Spiked Amount	50.000	Range 75 - 124	Recovery	= 100.040%		
50) Toluene-d8	8.653	98	336779	50.950	ug/l	0.00
Spiked Amount	50.000	Range 92 - 112	Recovery	= 101.900%		
62) 4-Bromofluorobenzene	11.085	95	107475	47.816	ug/l	0.00
Spiked Amount	50.000	Range 83 - 123	Recovery	= 95.640%		
Target Compounds						
					Qvalue	
16) Acetone	2.386	43	2431	2.318	ug/l #	88
20) Methylene Chloride	2.794	84	736	0.361	ug/l #	85
27) cis-1,2-Dichloroethene	4.495	96	758	0.365	ug/l	94
44) Trichloroethene	7.129	130	1147	0.541	ug/l	100
64) Tetrachloroethene	9.281	164	2519	1.285	ug/l	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX072221\
Data File : VX023401.D
Acq On : 23 Jul 2021 02:51
Operator : JC/MD
Sample : M3145-22
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 39 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
152140-FD-3

Quant Time: Jul 23 06:13:05 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X072221W.M
Quant Title : SW846 8260
QLast Update : Fri Jul 23 06:05:38 2021
Response via : Initial Calibration

