

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX072221\
 Data File : VX023402.D
 Acq On : 23 Jul 2021 03:14
 Operator : JC/MD
 Sample : M3146-06
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 40 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 152140-MW-E

Quant Time: Jul 23 06:13:17 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X072221W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jul 23 06:05:38 2021
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	172669	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	291577	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.061	117	249735	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	94313	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	112459	52.212	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	= 104.420%	
35) Dibromofluoromethane	5.397	113	93522	51.299	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 102.600%	
50) Toluene-d8	8.653	98	342440	51.126	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	= 102.260%	
62) 4-Bromofluorobenzene	11.085	95	105078	46.136	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 92.280%	
Target Compounds						
					Qvalue	
16) Acetone	2.386	43	3364	3.184	ug/l #	82
20) Methylene Chloride	2.794	84	982	0.478	ug/l #	89
44) Trichloroethene	7.129	130	2361	1.099	ug/l	92
64) Tetrachloroethene	9.281	164	17203	8.655	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX072221\
Data File : VX023402.D
Acq On : 23 Jul 2021 03:14
Operator : JC/MD
Sample : M3146-06
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 40 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampled :
152140-MW-E

Quant Time: Jul 23 06:13:17 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X072221W.M
Quant Title : SW846 8260
QLast Update : Fri Jul 23 06:05:38 2021
Response via : Initial Calibration

