

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX072221\
 Data File : VX023404.D
 Acq On : 23 Jul 2021 04:01
 Operator : JC/MD
 Sample : PB137858ZHE#05
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 42 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB137858ZHE#05

Quant Time: Jul 23 06:13:37 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X072221W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jul 23 06:05:38 2021
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	176043	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	301958	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	260299	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	94311	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	114332	52.064	ug/l	0.00
Spiked Amount 50.000	Range 78 - 117		Recovery	= 104.120%		
35) Dibromofluoromethane	5.397	113	94956	50.295	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery	= 100.600%		
50) Toluene-d8	8.653	98	358993	51.754	ug/l	0.00
Spiked Amount 50.000	Range 92 - 112		Recovery	= 103.500%		
62) 4-Bromofluorobenzene	11.085	95	108359	45.940	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery	= 91.880%		
Target Compounds						Qvalue
16) Acetone	2.386	43	31145	28.912	ug/l	92
20) Methylene Chloride	2.794	84	16138	7.702	ug/l	92
43) Isopropyl Acetate	6.355	43	25924	4.858	ug/l	92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX072221\
Data File : VX023404.D
Acq On : 23 Jul 2021 04:01
Operator : JC/MD
Sample : PB137858ZHE#05
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 42 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB137858ZHE#05

Quant Time: Jul 23 06:13:37 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X072221W.M
Quant Title : SW846 8260
QLast Update : Fri Jul 23 06:05:38 2021
Response via : Initial Calibration

