

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX072221\
 Data File : VX023406.D
 Acq On : 23 Jul 2021 04:49
 Operator : JC/MD
 Sample : PB137858ZHE#07
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 44 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB137858ZHE#07

Quant Time: Jul 23 06:13:57 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X072221W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jul 23 06:05:38 2021
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	177920	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	304073	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	278087	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	111744	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	115760	52.158	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	= 104.320%	
35) Dibromofluoromethane	5.397	113	93991	49.438	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 98.880%	
50) Toluene-d8	8.653	98	367445	52.604	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	= 105.200%	
62) 4-Bromofluorobenzene	11.085	95	119437	50.285	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 100.560%	
Target Compounds						Qvalue
16) Acetone	2.386	43	30557	28.067	ug/l	95
20) Methylene Chloride	2.794	84	15719	7.423	ug/l	90
43) Isopropyl Acetate	6.354	43	25982	4.835	ug/l	92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX072221\
Data File : VX023406.D
Acq On : 23 Jul 2021 04:49
Operator : JC/MD
Sample : PB137858ZHE#07
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 44 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB137858ZHE#07

Quant Time: Jul 23 06:13:57 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X072221W.M
Quant Title : SW846 8260
QLast Update : Fri Jul 23 06:05:38 2021
Response via : Initial Calibration

