

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX072221\
 Data File : VX023407.D
 Acq On : 23 Jul 2021 05:12
 Operator : JC/MD
 Sample : M3128-02
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 45 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 ETGI-358

Quant Time: Jul 23 08:09:48 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X072221W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jul 23 06:05:38 2021
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	174813	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	291792	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	251031	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	89674	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	112879	51.764	ug/l	0.00
Spiked Amount 50.000	Range 78 - 117		Recovery	= 103.520%		
35) Dibromofluoromethane	5.397	113	92223	50.549	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery	= 101.100%		
50) Toluene-d8	8.653	98	348700	52.022	ug/l	0.00
Spiked Amount 50.000	Range 92 - 112		Recovery	= 104.040%		
62) 4-Bromofluorobenzene	11.085	95	103183	45.270	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery	= 90.540%		
Target Compounds						Qvalue
16) Acetone	2.386	43	58782	54.952	ug/l	92
20) Methylene Chloride	2.794	84	17426	8.375	ug/l	93
25) 2-Butanone	4.580	43	15457	9.308	ug/l	95
43) Isopropyl Acetate	6.355	43	23213	4.501	ug/l	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX072221\
Data File : VX023407.D
Acq On : 23 Jul 2021 05:12
Operator : JC/MD
Sample : M3128-02
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 45 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
ETGI-358

Quant Time: Jul 23 08:09:48 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X072221W.M
Quant Title : SW846 8260
QLast Update : Fri Jul 23 06:05:38 2021
Response via : Initial Calibration

