

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX072221\
 Data File : VX023408.D
 Acq On : 23 Jul 2021 05:36
 Operator : JC/MD
 Sample : M3129-02
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 46 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 RO-VNJ-207

Quant Time: Jul 23 08:10:10 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X072221W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jul 23 06:05:38 2021
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	176945	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	299880	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	265644	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	107267	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	115351	52.260	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	= 104.520%	
35) Dibromofluoromethane	5.397	113	94446	50.372	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 100.740%	
50) Toluene-d8	8.653	98	359109	52.130	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	= 104.260%	
62) 4-Bromofluorobenzene	11.085	95	112685	48.106	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 96.220%	
Target Compounds						Qvalue
16) Acetone	2.386	43	38141	35.226	ug/l	93
20) Methylene Chloride	2.794	84	16820	7.986	ug/l	88
43) Isopropyl Acetate	6.348	43	23640	4.460	ug/l	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX072221\
Data File : VX023408.D
Acq On : 23 Jul 2021 05:36
Operator : JC/MD
Sample : M3129-02
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 46 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
RO-VNJ-207

Quant Time: Jul 23 08:10:10 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X072221W.M
Quant Title : SW846 8260
QLast Update : Fri Jul 23 06:05:38 2021
Response via : Initial Calibration

