

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX072221\
 Data File : VX023387.D
 Acq On : 22 Jul 2021 20:57
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Jul 23 06:09:57 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X072221W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jul 23 06:05:38 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	50.000	50.000	0.0	95	0.00
2 T	Dichlorodifluoromethane	50.000	52.676	-5.4	94	0.00
3 P	Chloromethane	50.000	50.481	-1.0	93	0.00
4 C	Vinyl Chloride	50.000	51.978	-4.0#	94	0.00
5 T	Bromomethane	50.000	49.998	0.0	98	0.00
6 T	Chloroethane	50.000	53.669	-7.3	96	0.00
7 T	Trichlorofluoromethane	50.000	52.377	-4.8	95	0.00
8 T	Diethyl Ether	50.000	51.793	-3.6	94	0.00
9 T	1,1,2-Trichlorotrifluoroeth	50.000	51.610	-3.2	94	0.00
10 T	Methyl Iodide	50.000	52.510	-5.0	91	0.00
11 T	Tert butyl alcohol	250.000	254.962	-2.0	94	0.00
12 CM	1,1-Dichloroethene	50.000	51.559	-3.1#	95	0.00
13 T	Acrolein	250.000	274.625	-9.8	104	0.00
14 T	Allyl chloride	50.000	50.888	-1.8	93	0.00
15 T	Acrylonitrile	250.000	265.539	-6.2	94	0.00
16 T	Acetone	250.000	243.893	2.4	94	0.00
17 T	Carbon Disulfide	50.000	49.334	1.3	90	0.00
18 T	Methyl Acetate	50.000	53.065	-6.1	97	0.00
19 T	Methyl tert-butyl Ether	50.000	52.433	-4.9	94	0.00
20 T	Methylene Chloride	50.000	48.758	2.5	96	0.00
21 T	trans-1,2-Dichloroethene	50.000	51.378	-2.8	94	0.00
22 T	Diisopropyl ether	50.000	53.574	-7.1	96	0.00
23 T	Vinyl Acetate	250.000	269.341	-7.7	94	0.00
24 P	1,1-Dichloroethane	50.000	52.827	-5.7	95	0.00
25 T	2-Butanone	250.000	262.930	-5.2	94	0.00
26 T	2,2-Dichloropropane	50.000	46.725	6.5	84	0.00
27 T	cis-1,2-Dichloroethene	50.000	53.373	-6.7	97	0.00
28 T	Bromochloromethane	50.000	49.827	0.3	93	0.00
29 T	Tetrahydrofuran	250.000	268.841	-7.5	96	0.00
30 C	Chloroform	50.000	52.322	-4.6#	95	0.00
31 T	Cyclohexane	50.000	50.256	-0.5	94	0.00
32 T	1,1,1-Trichloroethane	50.000	53.671	-7.3	94	0.00
33 S	1,2-Dichloroethane-d4	50.000	50.185	-0.4	98	0.00
34 I	1,4-Difluorobenzene	50.000	50.000	0.0	96	0.00
35 S	Dibromofluoromethane	50.000	50.934	-1.9	99	0.00
36 T	1,1-Dichloropropene	50.000	51.948	-3.9	95	0.00
37 T	Ethyl Acetate	50.000	51.511	-3.0	96	0.00
38 T	Carbon Tetrachloride	50.000	52.031	-4.1	94	0.00
39 T	Methylcyclohexane	50.000	51.290	-2.6	93	0.00
40 TM	Benzene	50.000	52.439	-4.9	96	0.00
41 T	Methacrylonitrile	50.000	51.895	-3.8	96	0.00
42 TM	1,2-Dichloroethane	50.000	51.959	-3.9	95	0.00
43 T	Isopropyl Acetate	50.000	52.507	-5.0	94	0.00
44 TM	Trichloroethene	50.000	52.478	-5.0	96	0.00
45 C	1,2-Dichloropropane	50.000	52.479	-5.0#	95	0.00
46 T	Dibromomethane	50.000	53.122	-6.2	96	0.00
47 T	Bromodichloromethane	50.000	53.649	-7.3	94	0.00
48 T	Methyl methacrylate	50.000	53.125	-6.3	94	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX072221\
 Data File : VX023387.D
 Acq On : 22 Jul 2021 20:57
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Jul 23 06:09:57 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X072221W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jul 23 06:05:38 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	1000.000	1023.153	-2.3	94	0.00
50 S	Toluene-d8	50.000	50.255	-0.5	97	0.00
51 T	4-Methyl-2-Pentanone	250.000	268.035	-7.2	94	0.00
52 CM	Toluene	50.000	52.845	-5.7#	95	0.00
53 T	t-1,3-Dichloropropene	50.000	47.574	4.9	90	0.00
54 T	cis-1,3-Dichloropropene	50.000	52.458	-4.9	91	0.00
55 T	1,1,2-Trichloroethane	50.000	54.668	-9.3	95	0.00
56 T	Ethyl methacrylate	50.000	54.062	-8.1	93	0.00
57 T	1,3-Dichloropropane	50.000	53.070	-6.1	95	0.00
58 T	2-Chloroethyl Vinyl ether	250.000	259.086	-3.6	95	0.00
59 T	2-Hexanone	250.000	271.254	-8.5	95	0.00
60 T	Dibromochloromethane	50.000	48.789	2.4	93	0.00
61 T	1,2-Dibromoethane	50.000	53.895	-7.8	94	0.00
62 S	4-Bromofluorobenzene	50.000	50.636	-1.3	98	0.00
63 I	Chlorobenzene-d5	50.000	50.000	0.0	94	0.00
64 T	Tetrachloroethene	50.000	51.945	-3.9	94	0.00
65 PM	Chlorobenzene	50.000	51.676	-3.4	93	0.00
66 T	1,1,1,2-Tetrachloroethane	50.000	49.888	0.2	92	0.00
67 C	Ethyl Benzene	50.000	53.055	-6.1#	94	0.00
68 T	m/p-Xylenes	100.000	104.600	-4.6	92	0.00
69 T	o-Xylene	50.000	52.273	-4.5	94	0.00
70 T	Styrene	50.000	54.360	-8.7	95	0.00
71 P	Bromoform	50.000	46.625	6.8	91	0.00
72 I	1,4-Dichlorobenzene-d4	50.000	50.000	0.0	95	0.00
73 T	Isopropylbenzene	50.000	52.706	-5.4	93	0.00
74 T	N-ethyl acetate	50.000	53.637	-7.3	93	0.00
75 P	1,1,2,2-Tetrachloroethane	50.000	52.111	-4.2	95	0.00
76 T	1,2,3-Trichloropropane	50.000	53.963	-7.9	95	0.00
77 T	Bromobenzene	50.000	54.000	-8.0	95	0.00
78 T	n-propylbenzene	50.000	53.236	-6.5	92	0.00
79 T	2-Chlorotoluene	50.000	52.643	-5.3	93	0.00
80 T	1,3,5-Trimethylbenzene	50.000	53.244	-6.5	93	0.00
81 T	trans-1,4-Dichloro-2-butene	50.000	50.519	-1.0	87	0.00
82 T	4-Chlorotoluene	50.000	52.348	-4.7	93	0.00
83 T	tert-Butylbenzene	50.000	53.492	-7.0	94	0.00
84 T	1,2,4-Trimethylbenzene	50.000	53.390	-6.8	94	0.00
85 T	sec-Butylbenzene	50.000	53.224	-6.4	95	0.00
86 T	p-Isopropyltoluene	50.000	53.879	-7.8	94	0.00
87 T	1,3-Dichlorobenzene	50.000	52.259	-4.5	94	0.00
88 T	1,4-Dichlorobenzene	50.000	51.305	-2.6	95	0.00
89 T	n-Butylbenzene	50.000	51.084	-2.2	91	0.00
90 T	Hexachloroethane	50.000	47.927	4.1	91	0.00
91 T	1,2-Dichlorobenzene	50.000	52.245	-4.5	96	0.00
92 T	1,2-Dibromo-3-Chloropropane	50.000	48.105	3.8	92	0.00
93 T	1,2,4-Trichlorobenzene	50.000	53.033	-6.1	94	0.00
94 T	Hexachlorobutadiene	50.000	50.524	-1.0	94	0.00
95 T	Naphthalene	50.000	55.286	-10.6	94	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX072221\
Data File : VX023387.D
Acq On : 22 Jul 2021 20:57
Operator : JC/MD
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 24 Sample Multiplier: 1

Instrument :
MSVOA_X
LabSampleId :
VSTDCCC050

Quant Time: Jul 23 06:09:57 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X072221W.M
Quant Title : SW846 8260
QLast Update : Fri Jul 23 06:05:38 2021
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	50.000	53.184	-6.4	94	0.00

(#) = Out of Range SPCC's out = 0 CCC's out = 6