

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX072222\
 Data File : VX030180.D
 Acq On : 22 Jul 2022 15:15
 Operator : JC/MD
 Sample : PB146364TB
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB146364TB

Quant Time: Jul 22 23:50:44 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X071222W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jul 13 09:05:23 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	173852	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	323710	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	286338	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	122651	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	115995	54.816	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery = 109.640%			
35) Dibromofluoromethane	5.385	113	96598	46.711	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery = 93.420%			
50) Toluene-d8	8.653	98	378310	48.641	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery = 97.280%			
62) 4-Bromofluorobenzene	11.085	95	128433	47.585	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery = 95.160%			
Target Compounds					Qvalue	
16) Acetone	2.385	43	34354	34.453	ug/l	91
20) Methylene Chloride	2.794	84	10622	3.514	ug/l	97
25) 2-Butanone	4.574	43	11513	6.778	ug/l	92
37) Ethyl Acetate	4.720	43	46282	13.484	ug/l	99
43) Isopropyl Acetate	6.348	43	188610	36.543	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX072222\
Data File : VX030180.D
Acq On : 22 Jul 2022 15:15
Operator : JC/MD
Sample : PB146364TB
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB146364TB

Quant Time: Jul 22 23:50:44 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X071222W.M
Quant Title : SW846 8260
QLast Update : Wed Jul 13 09:05:23 2022
Response via : Initial Calibration

