

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX072318\
 Data File : VX003600.D
 Acq On : 23 Jul 2018 13:23
 Operator : JC/SP
 Sample : J4048-12
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TB-2018.07.17

Quant Time: Jul 24 10:13:39 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X071918W.M
 Quant Title : SW846 8260
 QLast Update : Thu Jul 19 17:09:44 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	253072	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.87	114	378997	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	360999	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	196735	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	161513	53.92	ug/l	0.00
Spiked Amount	50.000		Recovery	=	107.84%	
35) Dibromofluoromethane	5.51	113	144751	50.14	ug/l	0.00
Spiked Amount	50.000		Recovery	=	100.28%	
50) Toluene-d8	8.72	98	541123	49.87	ug/l	0.00
Spiked Amount	50.000		Recovery	=	99.74%	
62) 4-Bromofluorobenzene	11.14	95	182115	45.99	ug/l	0.00
Spiked Amount	50.000		Recovery	=	91.98%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX072318\
Data File : VX003600.D
Acq On : 23 Jul 2018 13:23
Operator : JC/SP
Sample : J4048-12
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TB-2018.07.17

Quant Time: Jul 24 10:13:39 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X071918W.M
Quant Title : SW846 8260
QLast Update : Thu Jul 19 17:09:44 2018
Response via : Initial Calibration

