

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX072318\
 Data File : VX003602.D
 Acq On : 23 Jul 2018 14:16
 Operator : JC/SP
 Sample : J4083-01
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MH-6

Quant Time: Jul 24 10:14:42 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X071918W.M
 Quant Title : SW846 8260
 QLast Update : Thu Jul 19 17:09:44 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	243767	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.87	114	361891	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	348123	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	186934	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	157805	54.69	ug/l	0.00
Spiked Amount 50.000			Recovery	=	109.38%	
35) Dibromofluoromethane	5.51	113	142594	51.73	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.46%	
50) Toluene-d8	8.72	98	517871	49.99	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.98%	
62) 4-Bromofluorobenzene	11.14	95	176583	46.70	ug/l	0.00
Spiked Amount 50.000			Recovery	=	93.40%	
Target Compounds						
8) Diethyl Ether	2.19	74	7926	4.47	ug/l	95
16) Acetone	2.45	43	7778	7.14	ug/l #	89
29) Tetrahydrofuran	5.18	42	3516	2.91	ug/l #	93
40) Benzene	6.14	78	3675	0.31	ug/l	96
65) Chlorobenzene	10.14	112	10792	1.20	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX072318\
Data File : VX003602.D
Acq On : 23 Jul 2018 14:16
Operator : JC/SP
Sample : J4083-01
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MH-6

Quant Time: Jul 24 10:14:42 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X071918W.M
Quant Title : SW846 8260
QLast Update : Thu Jul 19 17:09:44 2018
Response via : Initial Calibration

