

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX072318\
 Data File : VX003604.D
 Acq On : 23 Jul 2018 15:09
 Operator : JC/SP
 Sample : J4059-01
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 DSN001

Quant Time: Jul 24 07:24:03 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X071918W.M
 Quant Title : SW846 8260
 QLast Update : Thu Jul 19 17:09:44 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	252603	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.87	114	379593	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	364418	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	201595	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	6.07	65	163853	54.80	ug/l	0.00
Spiked Amount 50.000			Recovery	=	109.60%	
35) Dibromofluoromethane	5.51	113	145769	50.42	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.84%	
50) Toluene-d8	8.72	98	548031	50.43	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.86%	
62) 4-Bromofluorobenzene	11.14	95	187277	47.22	ug/l	0.00
Spiked Amount 50.000			Recovery	=	94.44%	

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
16) Acetone	2.45	43	16765	14.84	ug/l	# 86

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX072318\
Data File : VX003604.D
Acq On : 23 Jul 2018 15:09
Operator : JC/SP
Sample : J4059-01
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
DSN001

Quant Time: Jul 24 07:24:03 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X071918W.M
Quant Title : SW846 8260
QLast Update : Thu Jul 19 17:09:44 2018
Response via : Initial Calibration

